



# PHYSICAL REFERENCE STANDARDS

Paragon Scientific  Ltd

**Industrial**  
VHG | ARMI | MBH

LGC Quality | ISO 17034 | ISO/IEC 17025 | ISO 9001

[lgcstandards.com/industrial](http://lgcstandards.com/industrial) | [InfoUSA@LGCGroup.com](mailto:InfoUSA@LGCGroup.com) | [paragon.sales@lgcgroup.com](mailto:paragon.sales@lgcgroup.com)

**Science for a Safer World**

# CONTENTS

## Physical Reference Standards

<b>Introduction</b>	<b>4</b>	<b>Section 4: Sucrose Brix Reference Standards</b>	<b>45</b>	<b>Section 13: Conductivity Reference Standards</b>	<b>59</b>
<b>Section 1: Viscosity Reference Standards</b>	<b>5</b>	Paragon Scientific Sucrose Brix/RI Reference Standards	45	Paragon Scientific Conductivity Reference Standards	59
Paragon Scientific General Purpose Reference Standards	7	<b>Section 5: Smoke Point Reference Standards</b>	<b>46</b>	<b>Section 14: Colour Reference Standards</b>	<b>60</b>
Paragon Scientific Cold Crank Simulator Reference Standards	14	Paragon Scientific Smoke Point Reference Standards	46	Paragon Scientific Colour Reference Standards – ASTM Method	60
Paragon Scientific Cone & Plate Reference Standards	17	<b>Section 6: Relative Density Reference Standards</b>	<b>47</b>	Paragon Scientific Colour Reference Standards – Saybolt Method	60
Paragon Scientific Flow Cup Reference Standards	22	Paragon Scientific Relative Density Reference Standards	47	Paragon Scientific Colour Reference Standards – Lovibond RYBN Method	61
Paragon Scientific High Temperature Reference Standards	24	<b>Section 7: Multi-Parameter Reference Standards</b>	<b>48</b>	Paragon Scientific Colour Reference Standards – Pt-Co/Hazen/APHA	61
Paragon Scientific Low Temperature Reference Standards	26	Paragon Scientific Multi-Parameter Reference Standards	48	Paragon Scientific Colour Reference Standards – Gardner Method	62
Paragon Scientific Medical Grade Reference Standards	29	<b>Section 8: Refractive Index Reference Standards</b>	<b>49</b>	Paragon Scientific Colour Reference Standards – AOCS-Tintometer Method	62
Paragon Scientific Mineral Oil Rotational Reference Standards	31	Paragon Scientific Refractive Index Reference Standards	49	<b>Section 15: Simulated Distillation Reference Standards</b>	<b>63</b>
Paragon Scientific Rotational Type Reference Standards	34	<b>Section 9: Flash Point Reference Standards</b>	<b>50</b>	VHG Simulated Distillation Reference Standards	63
Paragon Scientific Bath Media Reference Standards	36	Paragon Scientific Cleveland Open Cup Method Reference Standards	50	<b>Section 16: Red Eye in Diesel Reference Standards</b>	<b>63</b>
Paragon Scientific Check Oil Reference Standards	36	Paragon Scientific Pensky Martens Method Reference Standards	51	VHG Red Dye in Diesel Reference Standards	63
Paragon Scientific Small Sample Reference Standards	37	<b>Section 10: Certified Ethanol Reference Standards</b>	<b>51</b>	<b>Section 17: Soot in Diesel Reference Standards</b>	<b>64</b>
VHG Viscosity Reference Standards	39	Paragon Scientific Certified Ethanol Reference Standards	51	VHG Soot in Diesel Reference Standards	64
Pure Water Viscosity Reference Standards	40	<b>Section 11: Density Reference Standards</b>	<b>52</b>	<b>Section 18: Moisture Content Reference Standards</b>	<b>65</b>
<b>Section 2: Base Number Reference Standards</b>	<b>41</b>	Paragon Scientific Pure Water Density Reference Standards	52	VHG Crackle Test Reference Standards	65
Paragon Scientific Base Number Reference Standards	41	Paragon Scientific Density Reference Standards	52	VHG Titration Reference Standards	65
VHG Base Number Reference Standards	42	<b>Section 12: Single Parameter Certified Reference Materials (CRMs and CRMUs)</b>	<b>55</b>	<b>Section 19: Particle Count Reference Standards</b>	<b>66</b>
<b>Section 3: Acid Number Reference Standards</b>	<b>43</b>	Paragon Scientific Certified Reference Materials under UKAS Accreditation Standards	55	VHG Particle Count Reference Standards	66
Paragon Scientific Acid Number Reference Standards	43	Paragon Scientific Certified Reference Materials	57		
VHG Acid Number Reference Standards	44				



# Introduction

For over 25 years our Paragon Scientific site has specialised in and continues to produce a comprehensive range of certified reference materials (CRMs) and standards. Paragon holds dual UKAS accreditations ISO 17025 & ISO 17034, ensuring the highest level of product quality, metrology and traceability. We offer one of the broadest ranges of viscosity CRMs in the industry, as well as a wide range of other physical property standards. Our customers include some of the largest petrochemical and fuel

testing laboratories in the world, but our products are also used by many food, environmental, cannabis and pharmaceutical testing laboratories.

---

With nearly 50 years of experience the LGC Industrials portfolio is comprised of industry leading brands including VHG, ARMI, MBH and Paragon Scientific. Building on our collective expertise, Industrial fosters the innovation and the agility necessary to create the measurement tools you rely on when developing, using, and transforming materials to achieve your mission.

**LGC Industrial – The Material Difference.**

ISO/IEC 17043



## Section 1

# Viscosity Reference Standards

### Viscosity

LGC Industrial's produce dual accredited ISO 17025 / ISO 17034 viscosity standards for the calibration and verification of all types of viscometers, where operating temperature is controlled precisely. All viscosity standards are manufactured from high quality raw materials.

### General Purpose Viscosity Reference Standards

General Purpose Viscosity Standards are certified at all temperatures in strict accordance with the primary method for the certification of oil-based viscosity standards ASTM D2162 and ASTM D1480. Data quoted on certificate for Kinematic Viscosity, Dynamic Viscosity and Density at a range of temperatures between 20 and 100 °C

All General Purpose standards are applicable for use with ASTM D445, ASTM D446, IP 71 Section 1, IP 71 Section 2, ISO 3104, ISO 3105, and other internationally equivalent methodology.

### Cold Crank Simulator (CCS) Viscosity Reference Standards

Cold Crank Simulator (CCS) Viscosity Standards are for the calibration and verification of analytical equipment used in (CCS) oil testing to ASTM D5293, "SAE Specification J300" and other internationally equivalent methodology. Dynamic viscosity (mPa·s) and kinematic viscosity mm<sup>2</sup>/s (cSt) data is quoted on the certificate.

Can also be used for other low temperature applications.

### Low Temperature Viscosity Reference Standards

Low Temperature Viscosity Standards are widely used for the calibration and verification of viscometers used at sub-zero temperature. Low Temperature Viscosity Standards are applicable for use with ASTM D445, D2983, D3829, D4684, and other internationally equivalent methodology.

### High Temperature Viscosity Reference Standards

High Temperature Viscosity Standards offer additional certified data to the General Purpose Viscosity Standards, up to a temperature of 150 °C. All High Temperature Viscosity Standards are applicable for use with ASTM D445, ASTM D446, IP 71 Section 1, IP 71 Section 2, ISO 3104, ISO 3105, and other internationally equivalent methodology.

### Small Sample Viscosity Reference Standards

The Small Sample Viscosity Standards offer a cost-effective 60ml volume, reducing potential waste for the low volume viscometer user. Standards are certified for kinematic viscosity mm<sup>2</sup>/s (cSt) at 40 and 100 °C, in strict accordance with ASTM D2162.

### Rotational Viscosity Reference Standards

We offer a wide range of Rotational Viscosity Standards in two matrices; Mineral Oil and Silicon Fluid. These are applicable across a wide range of industries.

## Section 1: Viscosity Reference Standards

### Silicone Rotational Viscosity Reference Standards

Silicone Rotational Viscosity Standards are specifically formulated for calibration and verification of rotational viscometers. They are certified at 20 and 25 °C, for Kinematic Viscosity mm<sup>2</sup>/s (cSt), Dynamic Viscosity (mPa·s) and Density (g/mL).

### Mineral Oil Rotational Viscosity Reference Standards

Mineral Oil Rotational Viscosity Standards are the first choice option where end users are unable to have silicone in their process. They are certified for dynamic viscosity from 20 °C through to 25 °C at 0.5 °C intervals. The dynamic viscosities at 20 and 25 °C were derived from the kinematic viscosities certified in strict accordance with ASTM D2162 and the densities certified in strict accordance with ASTM D1480. Dynamic viscosities at intermediate temperatures were derived from the kinematic viscosities calculated using ASTM D341 and densities calculated by proportional calculations.

### Flow Cup Viscosity Reference Standards

Flow Cup Viscosity Standards are designed for use with most flow cup viscometers. Certified at 20 and 25 °C for Kinematic Viscosity mm<sup>2</sup>/s (cSt), Dynamic Viscosity (mPa·s) and Density (g/mL), they are also supplied with calculated drain times, for the Ford, ISO, Shell, and Zahn flow cups, derived from the kinematic viscosity and calculations as defined in the methodology. A table of flow cups and calculations is also included on the reverse of the certificate. Applicable methods include, but are not limited to, ASTM D1200, D4212, ISO 2431 and other internationally equivalent methodology.

### Cone and Plate Viscosity Reference Standards

Our Cone and Plate Viscosity Standards are specifically formulated for viscometers used in the paints and coatings industry. Certified at 20 and 25 °C, quoting Kinematic Viscosity mm<sup>2</sup>/s (cSt), Dynamic Viscosity (mPa·s) and Density (g/mL). Where applicable, Krebs Units (KU) will also be quoted.

### Water Viscosity Reference Standards

Our Pure Water Viscosity Standard is certified for Kinematic Viscosity mm<sup>2</sup>/s (cSt), Dynamic Viscosity (mPa·s) and Density (g/mL), at 5, 20, 25 and 37 °C. Certification is carried out in strict accordance with ASTM D445 at 5 °C and ASTM D2162 at 20, 25 & 37 °C. Density determination is carried out at all temperatures, in strict accordance with "ASTM D1480.

### Medical Viscosity Reference Standards

We produce medical grade viscosity standards which are water / glycerine based. Certified for Kinematic Viscosity mm<sup>2</sup>/s (cSt), Dynamic Viscosity (mPa·s) and Density (g/mL), at 25 and 37 °C in strict accordance with ASTM D2162 and ASTM D1480.

### Custom Blend Viscosity Reference Standards

If a customer has a requirement for a different temperature or viscosity value other than those quoted, please contact the Paragon technical department for further information. At Paragon Scientific we take pride in making customer needs our priority and we have a long history of supplying custom blends to many satisfied end-users.



## Section 1: Viscosity Reference Standards

Paragon Scientific – Viscosity Reference Standards – General Purpose		
Product Number	Product Description	Size (ml)
<a href="#">ALK-D500</a>	Viscosity Reference Standard General Purpose Type D500	500ml
<a href="#">ALK-D5000</a>	Viscosity Reference Standard General Purpose Type D5000	500ml
<a href="#">ALK-D7500</a>	Viscosity Reference Standard General Purpose Type D7500	500ml
<a href="#">ALK-N.4</a>	Viscosity Reference Standard General Purpose Type N.4	500ml
<a href="#">ALK-N.8</a>	Viscosity Reference Standard General Purpose Type N.8	500ml
<a href="#">ALK-N1.0</a>	Viscosity Reference Standard General Purpose Type N1.0	500ml
<a href="#">ALK-N10</a>	Viscosity Reference Standard General Purpose Type N10	500ml
<a href="#">ALK-N100</a>	Viscosity Reference Standard General Purpose Type N100	500ml
<a href="#">ALK-N1000</a>	Viscosity Reference Standard General Purpose Type N1000	500ml
<a href="#">ALK-N10200</a>	Viscosity Reference Standard General Purpose Type N10200	500ml
<a href="#">ALK-N14</a>	Viscosity Reference Standard General Purpose Type N14	500ml
<a href="#">ALK-N140</a>	Viscosity Reference Standard General Purpose Type N140	500ml
<a href="#">ALK-N1400</a>	Viscosity Reference Standard General Purpose Type N1400	500ml
<a href="#">ALK-N15000</a>	Viscosity Reference Standard General Purpose Type N15000	500ml
<a href="#">ALK-N18000</a>	Viscosity Reference Standard General Purpose Type N18000	500ml
<a href="#">ALK-N2</a>	Viscosity Reference Standard General Purpose Type N2	500ml
<a href="#">ALK-N250</a>	Viscosity Reference Standard General Purpose Type N250	500ml
<a href="#">ALK-N2500</a>	Viscosity Reference Standard General Purpose Type N2500	500ml
<a href="#">ALK-N26</a>	Viscosity Reference Standard General Purpose Type N26	500ml
<a href="#">ALK-N35</a>	Viscosity Reference Standard General Purpose Type N35	500ml
<a href="#">ALK-N350</a>	Viscosity Reference Standard General Purpose Type N350	500ml
<a href="#">ALK-N4000</a>	Viscosity Reference Standard General Purpose Type N4000	500ml
<a href="#">ALK-N415</a>	Viscosity Reference Standard General Purpose Type N415	500ml
<a href="#">ALK-N44</a>	Viscosity Reference Standard General Purpose Type N44	500ml
<a href="#">ALK-N5100</a>	Viscosity Reference Standard General Purpose Type N5100	500ml
<a href="#">ALK-N7.5</a>	Viscosity Reference Standard General Purpose Type N7.5	500ml
<a href="#">ALK-N75</a>	Viscosity Reference Standard General Purpose Type N75	500ml
<a href="#">ALK-N750</a>	Viscosity Reference Standard General Purpose Type N750	500ml
<a href="#">ALK-S20</a>	Viscosity Reference Standard General Purpose Type S20	500ml
<a href="#">ALK-S200</a>	Viscosity Reference Standard General Purpose Type S200	500ml

Continued on the next page

## Section 1: Viscosity Reference Standards

Continued from previous page

Paragon Scientific – Viscosity Reference Standards – General Purpose		
Product Number	Product Description	Size (ml)
<a href="#">ALK-S2000</a>	Viscosity Reference Standard General Purpose Type S2000	500ml
<a href="#">ALK-S3</a>	Viscosity Reference Standard General Purpose Type S3	500ml
<a href="#">ALK-S30000</a>	Viscosity Reference Standard General Purpose Type S30000	500ml
<a href="#">ALK-S6</a>	Viscosity Reference Standard General Purpose Type S6	500ml
<a href="#">ALK-S60</a>	Viscosity Reference Standard General Purpose Type S60	500ml
<a href="#">ALK-S600</a>	Viscosity Reference Standard General Purpose Type S600	500ml
<a href="#">ALK-S8000</a>	Viscosity Reference Standard General Purpose Type S8000	500ml

General Purpose Viscosity Reference Standards									
Part No.	Kinematic Viscosity mm <sup>2</sup> /s (cSt)								
	20.00 °C	25 °C	37.78 °C	40 °C	50 °C	60 °C	80 °C	98.89 °C	100 °C
	68.00 °F	77.00 °F	100.00 °F	104.00 °F	122.00 °F	140.00 °F	176.00 °F	210.00 °F	212.00 °F
<a href="#">ALK-N.4</a>	0.4691	0.4495	0.4051	0.3981	-	-	-	-	-
<a href="#">ALK-N.8</a>	0.7286	0.6903	0.6080	0.5954	-	-	-	-	-
<a href="#">ALK-N1.0</a>	1.302	1.213	1.025	0.9976	0.8884	-	-	-	-
<a href="#">ALK-N2</a>	2.984	2.687	2.110	2.031	1.726	1.490	-	-	-
<a href="#">ALK-S3</a>	4.601	4.058	3.055	2.920	2.417	2.041	1.522	1.207	1.192
<a href="#">ALK-D5</a>	6.497	5.623	4.060	3.856	3.113	2.573	1.860	1.439	1.420
<a href="#">ALK-S6</a>	10.46	8.8869	6.117	5.772	4.528	3.655	2.540	1.913	1.885
<a href="#">ALK-N7.5</a>	12.34	10.45	7.163	6.748	5.263	4.220	2.902	2.167	2.131
<a href="#">ALK-D10</a>	14.63	12.18	8.075	7.572	5.799	4.583	3.086	2.273	2.236
<a href="#">ALK-N10</a>	20.92	17.05	10.79	10.05	7.479	5.775	3.749	2.693	2.646
<a href="#">ALK-N14</a>	30.54	24.46	14.87	13.76	9.977	7.529	4.713	3.304	3.241
<a href="#">ALK-S20</a>	42.73	33.65	19.68	18.10	12.82	9.476	5.734	3.924	3.845
<a href="#">ALK-N26</a>	61.14	49.10	29.80	27.53	19.81	14.79	9.020	6.160	6.034
<a href="#">ALK-N35</a>	86.53	65.98	35.89	32.62	22.01	15.61	8.824	5.747	5.620
<a href="#">ALK-N44</a>	111.2	86.37	48.42	44.19	30.14	21.47	12.10	7.791	7.610
<a href="#">ALK-S60</a>	155.3	115.2	58.80	52.93	34.24	23.41	12.47	7.773	7.582

Continued on the next page



## Section 1: Viscosity Reference Standards

Continued from previous page

General Purpose Viscosity Reference Standards									
Part No.	Kinematic Viscosity mm <sup>2</sup> /s (cSt)								
	20.00 °C	25 °C	37.78 °C	40 °C	50 °C	60 °C	80 °C	98.89 °C	100 °C
	68.00 °F	77.00 °F	100.00 °F	104.00 °F	122.00 °F	140.00 °F	176.00 °F	210.00 °F	212.00 °F
<a href="#">ALK-N75</a>	203.9	155.5	83.66	75.81	50.25	34.86	18.79	11.70	11.41
<a href="#">ALK-N100</a>	329.6	237.2	112.9	100.5	62.05	40.68	20.23	11.99	11.67
<a href="#">ALK-N140</a>	409.6	305.9	156.8	141.1	90.53	60.96	31.22	18.68	18.17
<a href="#">ALK-S200</a>	653.2	459.4	206.4	182.1	108.0	68.32	31.97	18.11	17.58
<a href="#">ALK-N250</a>	779.6	572.7	283.0	252.8	158.2	104.0	51.12	29.56	28.71
<a href="#">ALK-D500</a>	834.5	582.0	256.6	225.7	132.0	82.48	37.77	21.05	20.41
<a href="#">ALK-N350</a>	1240	850.7	361.7	316.1	180.4	110.1	48.51	26.26	25.42
<a href="#">ALK-N415</a>	1340	967.8	459.4	408.2	248.7	159.9	75.39	42.24	40.96
<a href="#">ALK-D1000</a>	1717	1170	487.4	424.5	238.6	143.5	61.57	32.60	31.54
<a href="#">ALK-S600</a>	2161	1460	595.4	516.9	286.2	169.9	71.19	37.05	35.81
<a href="#">ALK-N750</a>	2746	1925	852.4	749.0	435.3	268.3	117.8	62.52	60.45
<a href="#">ALK-N1000</a>	4489	2980	1162	1001	534.9	306.9	120.8	59.67	57.51
<a href="#">ALK-N1400</a>	5608	3822	1586	1378	764.3	451.5	184.4	92.50	89.13
<a href="#">ALK-S2000</a>	8573	5566	2053	1752	899.4	497.2	182.9	85.90	82.53
<a href="#">ALK-D5000</a>	10245	6670	2463	2101	1079	593.8	216.9	100.7	96.74
<a href="#">ALK-N2500</a>	11658	7673	2905	2488	1293	719.0	265.2	123.1	118.2
<a href="#">ALK-D7500</a>	13686	8817	3192	2713	1368	742.7	264.0	119.9	114.9
<a href="#">ALK-N4000</a>	17322	11176	4042	3435	1729	933.3	327.2	146.5	140.4
<a href="#">ALK-N5100</a>	26904	17187	6050	5110	2516	1330	446.9	192.9	184.5
<a href="#">ALKS8000</a>	37745	23822	8136	6853	3311	1724	560.5	235.9	225.3
<a href="#">ALK-N10200</a>	54473	34453	11868	9992	4834	2504	805.9	333.2	318.1
<a href="#">ALK-N15000</a>	74506	46451	15463	12954	6120	3102	962.0	385.9	367.7
<a href="#">ALK-N18000</a>	103824	65017	21802	18279	8654	4391	1352	537.3	511.3
<a href="#">ALK-S30000</a>	138056	85267	27723	23124	10717	5334	1586	612.3	581.5

## Section 1: Viscosity Reference Standards

General Purpose Viscosity Reference Standards									
Part No.	Dynamic Viscosity mPa·s (cP)								
	20.00 °C	25 °C	37.78 °C	40 °C	50 °C	60 °C	80 °C	98.89 °C	100 °C
	68.00 °F	77.00 °F	100.00 °F	104.00 °F	122.00 °F	140.00 °F	176.00 °F	210.00 °F	212.00 °F
<a href="#">ALK-N.4</a>	0.3094	0.2944	0.2606	0.2553	-	-	-	-	-
<a href="#">ALK-N.8</a>	0.5188	0.4887	0.4240	0.4142	-	-	-	-	-
<a href="#">ALK-N1.0</a>	1.014	0.9400	0.7843	0.7617	0.6715	-	-	-	-
<a href="#">ALK-N2</a>	2.427	2.176	1.689	1.623	1.367	1.170	-	-	-
<a href="#">ALK-S3</a>	3.780	3.320	2.472	2.358	1.936	1.620	1.187	0.9255	0.9131
<a href="#">ALK-D5</a>	5.446	4.695	3.355	3.180	2.546	2.087	1.483	1.129	1.113
<a href="#">ALK-S6</a>	8.703	7.350	5.018	4.725	3.677	2.943	2.012	1.491	1.467
<a href="#">ALK-N7.5</a>	10.09	8.514	5.776	5.431	4.201	3.341	2.259	1.659	1.63
<a href="#">ALK-D10</a>	12.17	10.09	6.621	6.198	4.709	3.691	2.445	1.772	1.742
<a href="#">ALK-N10</a>	17.66	14.34	8.986	8.355	6.168	4.726	3.019	2.135	2.096
<a href="#">ALK-N14</a>	25.69	20.50	12.34	11.40	8.200	6.140	3.783	2.612	2.56
<a href="#">ALK-S20</a>	36.11	28.33	16.41	15.07	10.59	7.767	4.627	3.120	3.054
<a href="#">ALK-N26</a>	49.98	39.98	24.02	22.16	15.82	11.72	7.031	4.728	4.627
<a href="#">ALK-N35</a>	74.13	56.31	30.34	27.53	18.44	12.98	7.228	4.640	4.534
<a href="#">ALK-N44</a>	92.04	71.23	39.55	36.03	24.39	17.24	9.569	6.071	5.924
<a href="#">ALK-S60</a>	133.7	98.83	49.97	44.92	28.84	19.58	10.27	6.313	6.153
<a href="#">ALK-N75</a>	169.8	129.0	68.76	62.20	40.92	28.17	14.96	9.180	8.944
<a href="#">ALK-N100</a>	285.6	204.8	96.60	85.84	52.62	34.25	16.79	9.813	9.544
<a href="#">ALK-N140</a>	343.2	255.4	129.7	116.5	74.23	49.62	25.04	14.77	14.35
<a href="#">ALK-S200</a>	568.4	398.4	177.4	156.3	92.04	57.82	26.68	14.91	14.46
<a href="#">ALK-N250</a>	657.1	481.0	235.5	210.1	130.5	85.17	41.25	23.53	22.83
<a href="#">ALK-D500</a>	727.1	505.4	220.8	193.9	112.6	69.90	31.56	17.36	16.81
<a href="#">ALK-N350</a>	1083	740.3	312.0	272.3	154.3	93.54	40.65	21.71	21
<a href="#">ALK-N415</a>	1136	817.4	384.5	341.1	206.3	131.7	61.22	33.84	32.78
<a href="#">ALK-D1000</a>	1503	1021	421.6	366.6	204.7	122.3	51.74	27.04	26.14
<a href="#">ALK-S600</a>	1894	1275	515.7	447.1	245.8	144.9	59.91	30.77	29.72

Continued on the next page

## Section 1: Viscosity Reference Standards

Continued from previous page

General Purpose Viscosity Reference Standards									
Part No.	Dynamic Viscosity mPa·s (cP)								
	20.00 °C	25 °C	37.78 °C	40 °C	50 °C	60 °C	80 °C	98.89 °C	100 °C
	68.00 °F	77.00 °F	100.00 °F	104.00 °F	122.00 °F	140.00 °F	176.00 °F	210.00 °F	212.00 °F
<a href="#">ALK-N750</a>	2352	1643	721.3	632.8	365.2	223.6	96.81	50.70	48.98
<a href="#">ALK-N1000</a>	3951	2615	1011	869.7	461.6	263.1	102.2	49.83	47.99
<a href="#">ALK-N1400</a>	4853	3296	1356	1176	648.1	380.3	153.3	75.89	73.06
<a href="#">ALK-S2000</a>	7571	4899	1792	1527	779.0	427.8	155.3	72.03	69.14
<a href="#">ALK-D5000</a>	9065	5882	2154	1835	936.4	512.0	184.6	84.63	81.24
<a href="#">ALK-N2500</a>	10205	6695	2514	2150	1110	613.2	223.2	102.3	98.19
<a href="#">ALK-D7500</a>	12119	7782	2795	2371	1188	640.9	224.9	100.9	96.6
<a href="#">ALK-N4000</a>	15368	9884	3546	3009	1505	807.2	279.3	123.5	118.3
<a href="#">ALK-N5100</a>	23912	15228	5318	4486	2195	1153	382.5	163.1	155.9
<a href="#">ALKS8000</a>	33604	21142	7164	6026	2893	1497	480.5	199.8	190.7
<a href="#">ALK-N10200</a>	48617	30656	10477	8809	4236	2180	693.1	283.2	270.1
<a href="#">ALK-N15000</a>	66571	41379	13666	11434	5368	2704	828.3	328.4	312.7
<a href="#">ALK-N18000</a>	92933	58021	19306	16164	7607	3836	1167	458.3	435.8
<a href="#">ALK-S30000</a>	123698	76169	24576	20469	9429	4665	1370	522.8	496.1



## Section 1: Viscosity Reference Standards

General Purpose Viscosity Reference Standards									
Part No.	Density g/mL								
	20.00 °C	25 °C	37.78 °C	40 °C	50 °C	60 °C	80 °C	98.89 °C	100 °C
	68.00 °F	77.00 °F	100.00 °F	104.00 °F	122.00 °F	140.00 °F	176.00 °F	210.00 °F	212.00 °F
<a href="#">ALK-N.4</a>	0.6595	0.6549	0.6434	0.6413	-	-	-	-	-
<a href="#">ALK-N.8</a>	0.7121	0.7079	0.6974	0.6956	-	-	-	-	-
<a href="#">ALK-N1.0</a>	0.7787	0.7749	0.7652	0.7635	0.7558	-	-	-	-
<a href="#">ALK-N2</a>	0.8132	0.8097	0.8007	0.7991	0.792	0.7849	-	-	-
<a href="#">ALK-S3</a>	0.8216	0.8181	0.8093	0.8077	0.8008	0.7939	0.78	0.7668	0.766
<a href="#">ALK-D5</a>	0.8383	0.8349	0.8263	0.8248	0.818	0.8111	0.7975	0.7846	0.7838
<a href="#">ALK-S6</a>	0.832	0.8287	0.8202	0.8186	0.812	0.8053	0.792	0.7793	0.7785
<a href="#">ALK-N7.5</a>	0.818	0.8147	0.8063	0.8048	0.7982	0.7916	0.7784	0.7658	0.7651
<a href="#">ALK-D10</a>	0.8317	0.8284	0.8199	0.8185	0.812	0.8054	0.7923	0.7797	0.779
<a href="#">ALK-N10</a>	0.8443	0.8411	0.8328	0.8313	0.8247	0.8183	0.8052	0.7929	0.7922
<a href="#">ALK-N14</a>	0.8448	0.8415	0.8332	0.8318	0.8254	0.819	0.8062	0.794	0.7934
<a href="#">ALK-S20</a>	0.8451	0.8419	0.8338	0.8324	0.8261	0.8197	0.807	0.795	0.7942
<a href="#">ALK-N26</a>	0.8174	0.8143	0.8061	0.8048	0.7985	0.7921	0.7795	0.7675	0.7668
<a href="#">ALK-N35</a>	0.8567	0.8535	0.8455	0.8441	0.8379	0.8316	0.8191	0.8074	0.8067
<a href="#">ALK-N44</a>	0.8277	0.8247	0.8168	0.8154	0.8093	0.8031	0.7908	0.7792	0.7785
<a href="#">ALK-S60</a>	0.861	0.8579	0.8499	0.8486	0.8423	0.8362	0.8239	0.8122	0.8115
<a href="#">ALK-N75</a>	0.8327	0.8297	0.8219	0.8205	0.8144	0.8082	0.796	0.7846	0.7839
<a href="#">ALK-N100</a>	0.8664	0.8634	0.8556	0.8541	0.848	0.842	0.8301	0.8184	0.8178
<a href="#">ALK-N140</a>	0.838	0.835	0.8273	0.826	0.8199	0.8139	0.8019	0.7906	0.7899
<a href="#">ALK-S200</a>	0.8702	0.8672	0.8595	0.8582	0.8522	0.8463	0.8344	0.8231	0.8225
<a href="#">ALK-N250</a>	0.8429	0.8399	0.8323	0.8309	0.825	0.8189	0.807	0.7959	0.7952
<a href="#">ALK-D500</a>	0.8713	0.8683	0.8606	0.8593	0.8533	0.8475	0.8356	0.8245	0.8238
<a href="#">ALK-N350</a>	0.8732	0.8702	0.8627	0.8613	0.8554	0.8496	0.8379	0.8268	0.8262
<a href="#">ALK-N415</a>	0.8476	0.8446	0.837	0.8357	0.8297	0.8239	0.8121	0.8011	0.8004
<a href="#">ALK-D1000</a>	0.8755	0.8725	0.865	0.8637	0.8579	0.852	0.8403	0.8294	0.8288
<a href="#">ALK-S600</a>	0.8765	0.8736	0.8661	0.8649	0.859	0.8531	0.8416	0.8306	0.8299
<a href="#">ALK-N750</a>	0.8565	0.8536	0.8462	0.8449	0.839	0.8333	0.8218	0.8109	0.8102
<a href="#">ALK-N1000</a>	0.8802	0.8774	0.87	0.8688	0.863	0.8573	0.8459	0.8351	0.8345

Continued on the next page

## Section 1: Viscosity Reference Standards

Continued from previous page

General Purpose Viscosity Reference Standards									
Part No.	Density g/mL								
	20.00 °C	25 °C	37.78 °C	40 °C	50 °C	60 °C	80 °C	98.89 °C	100 °C
	68.00 °F	77.00 °F	100.00 °F	104.00 °F	122.00 °F	140.00 °F	176.00 °F	210.00 °F	212.00 °F
<a href="#">ALK-N1400</a>	0.8653	0.8624	0.855	0.8537	0.848	0.8424	0.8311	0.8204	0.8197
<a href="#">ALK-S2000</a>	0.8831	0.8802	0.8728	0.8718	0.8661	0.8604	0.8491	0.8385	0.8377
<a href="#">ALK-D5000</a>	0.8848	0.8819	0.8747	0.8734	0.8678	0.8622	0.851	0.8404	0.8398
<a href="#">ALK-N2500</a>	0.8754	0.8726	0.8654	0.8642	0.8586	0.8529	0.8418	0.8313	0.8307
<a href="#">ALK-D7500</a>	0.8855	0.8826	0.8755	0.8741	0.8685	0.8629	0.8518	0.8412	0.8407
<a href="#">ALK-N4000</a>	0.8872	0.8844	0.8772	0.8759	0.8704	0.8649	0.8537	0.8432	0.8427
<a href="#">ALK-N5100</a>	0.8888	0.886	0.879	0.8778	0.8723	0.8667	0.8558	0.8455	0.8448
<a href="#">ALKS8000</a>	0.8903	0.8875	0.8805	0.8793	0.8737	0.8683	0.8573	0.847	0.8464
<a href="#">ALK-N10200</a>	0.8925	0.8898	0.8828	0.8816	0.8762	0.8708	0.86	0.8498	0.8492
<a href="#">ALK-N15000</a>	0.8935	0.8908	0.8838	0.8827	0.8772	0.8718	0.861	0.8509	0.8503
<a href="#">ALK-N18000</a>	0.8951	0.8924	0.8855	0.8843	0.879	0.8736	0.8629	0.8529	0.8523
<a href="#">ALK-S30000</a>	0.896	0.8933	0.8865	0.8852	0.8798	0.8745	0.8639	0.8538	0.8532



## Section 1: Viscosity Reference Standards

Paragon Scientific – Viscosity Reference Standards – Cold Crank Simulator		
Product Number	Product Description	Size (ml)
<a href="#">ALK-CL080</a>	Viscosity Reference Standard Cold Crank Simulator Type CL080 (CL08)	500ml
<a href="#">ALK-CL090</a>	Viscosity Reference Standard Cold Crank Simulator Type CL090 (CL09)	500ml
<a href="#">ALK-CL100</a>	Viscosity Reference Standard Cold Crank Simulator Type CL100 (CL10)	500ml
<a href="#">ALK-CL110</a>	Viscosity Reference Standard Cold Crank Simulator Type CL110 (CL11)	500ml
<a href="#">ALK-CL120</a>	Viscosity Reference Standard Cold Crank Simulator Type CL120 (CL12)	500ml
<a href="#">ALK-CL130</a>	Viscosity Reference Standard Cold Crank Simulator Type CL130 (CL13)	500ml
<a href="#">ALK-CL140</a>	Viscosity Reference Standard Cold Crank Simulator Type CL140 (CL14)	500ml
<a href="#">ALK-CL150</a>	Viscosity Reference Standard Cold Crank Simulator Type CL150 (CL15)	500ml
<a href="#">ALK-CL160</a>	Viscosity Reference Standard Cold Crank Simulator Type CL160 (CL16)	500ml
<a href="#">ALK-CL170</a>	Viscosity Reference Standard Cold Crank Simulator Type CL170 (CL17)	500ml
<a href="#">ALK-CL190</a>	Viscosity Reference Standard Cold Crank Simulator Type CL190 (CL19)	500ml
<a href="#">ALK-CL200</a>	Viscosity Reference Standard Cold Crank Simulator Type CL200 (CL20)	500ml
<a href="#">ALK-CL220</a>	Viscosity Reference Standard Cold Crank Simulator Type CL220 (CL22)	500ml
<a href="#">ALK-CL240</a>	Viscosity Reference Standard Cold Crank Simulator Type CL240 (CL24)	500ml
<a href="#">ALK-CL250</a>	Viscosity Reference Standard Cold Crank Simulator Type CL250 (CL25)	500ml
<a href="#">ALK-CL260</a>	Viscosity Reference Standard Cold Crank Simulator Type CL260 (CL26)	500ml
<a href="#">ALK-CL280</a>	Viscosity Reference Standard Cold Crank Simulator Type CL280 (CL28)	500ml
<a href="#">ALK-CL300</a>	Viscosity Reference Standard Cold Crank Simulator Type CL300 (CL30)	500ml
<a href="#">ALK-CL320</a>	Viscosity Reference Standard Cold Crank Simulator Type CL320 (CL32)	500ml
<a href="#">ALK-CL340</a>	Viscosity Reference Standard Cold Crank Simulator Type CL340 (CL34)	500ml

Continued on the next page

## Section 1: Viscosity Reference Standards

Continued from previous page

Paragon Scientific – Viscosity Reference Standards – Cold Crank Simulator		
Product Number	Product Description	Size (ml)
<a href="#">ALK-CL380</a>	Viscosity Reference Standard Cold Crank Simulator Type CL380 (CL38)	500ml
<a href="#">ALK-CL420</a>	Viscosity Reference Standard Cold Crank Simulator Type CL420 (CL42)	500ml
<a href="#">ALK-CL480</a>	Viscosity Reference Standard Cold Crank Simulator Type CL480 (CL48)	500ml
<a href="#">ALK-CL530</a>	Viscosity Reference Standard Cold Crank Simulator Type CL530 (CL53)	500ml
<a href="#">ALK-CL600</a>	Viscosity Reference Standard Cold Crank Simulator Type CL600 (CL60)	500ml
<a href="#">ALK-CL680</a>	Viscosity Reference Standard Cold Crank Simulator Type CL680 (CL68)	500ml
<a href="#">ALK-CL740</a>	Viscosity Reference Standard Cold Crank Simulator Type CL740 (CL74)	500ml
<a href="#">ALK-VIS-CCS-14</a>	Viscosity Set, Cold Crank Simulator Type: CL10, CL12, CL14, CL16, CL19, CL22, CL25 (2) , CL28, CL32, CL38, CL48, CL60, CL68	14x500ml
<a href="#">ALK-VIS-CCS-18</a>	Viscosity Set, Cold Crank Simulator Type: CL08, CL09, CL10, CL11, CL12, CL13, CL14, CL15, CL16, CL19, CL22, CL25(x2), CL28, CL32, CL38, CL48 & CL60	18x500ml
<a href="#">ALK-VIS-CCS-8</a>	Viscosity Set, Cold Crank Simulator: Type CL14, CL19, CL22, CL25, CL28, CL32, CL48, CL68	8x500ml



## Section 1: Viscosity Reference Standards

Cold Crank Simulator Viscosity Reference Standards										
Part No.	Dynamic Viscosity mPa·s (cP)									Size
	-5 °C	-10 °C	-15 °C	-18 °C	-20 °C	-25 °C	-30 °C	-35 °C	-40 °C	
<a href="#">CL08 (CL080)</a>	-	-	-		-	-	-	852.9	1404	500 mL
<a href="#">CL09 (CL090)</a>	-	-	-		-	-	-	1140	1909	500 mL
<a href="#">CL10 (CL100)</a>	-	-	-		-	-	883.3	1431	2418	500 mL
<a href="#">CL11 (CL110)</a>	-	-	-		-	-	1035	1694	2905	500 mL
<a href="#">CL12 (CL120)</a>	-	-	-		-	770.4	1221	2016	3481	500 mL
<a href="#">CL13 (CL130)</a>	-	-	-		-	947.5	1518	2534	4434	500 mL
<a href="#">CL14 (CL140)</a>	-	-	-	683.7	808.9	1264	2054	3482	6226	500 mL
<a href="#">CL15 (CL150)</a>	-	-	-		974.1	1535	2520	4323	7837	500 mL
<a href="#">CL16 (CL160)</a>	-	-	-	1014	1212	1939	3232	5657	10459	500 mL
<a href="#">CL17 (CL170)</a>	-	-	871.8		1356	2183	3671	6468	-	500 mL
<a href="#">CL19 (CL190)</a>	-	-	1078	1409	1694	2762	4717	8463	16150	500 mL
<a href="#">CL20 (CL200)</a>	-	862.4	1324		2105	3483	6022	10990	21337	500 mL
<a href="#">CL22 (CL220)</a>	-	1030	1593	2107	2553	4260	7436	13729	-	500 mL
<a href="#">CL24 (CL240)</a>	-	1211	1889		3053	5141	9061	16936	-	500 mL
<a href="#">CL25 (CL250)</a>	-	1375	2156	2873	3503	5934	10550	19870	-	500 mL
<a href="#">CL26 (CL260)</a>	-	1680	2646		4329	7390	13242	-	-	500 mL
<a href="#">CL28 (CL280)</a>	-	2022	3219	4326	5311	9179	16682	-	-	500 mL
<a href="#">CL30 (CL300)</a>	-	2431	3889		6468	11227	20590	-	-	500 mL
<a href="#">CL32 (CL320)</a>	-	3022	4863	6590	8156	14316	-	-	-	500 mL
<a href="#">CL34 (CL340)</a>	-	3431	5567		9395	16648	-	-	-	500 mL
<a href="#">CL38 (CL380)</a>	2670	4199	6868	9400	11722	20950	-	-	-	500 mL
<a href="#">CL42 (CL420)</a>	-	4969	8161		13955	-	-	-	-	500 mL
<a href="#">CL48 (CL480)</a>	3840	6131	10151	14042	17610	-	-	-	-	500 mL
<a href="#">CL53 (CL530)</a>	-	7440	12466		21863	-	-	-	-	500 mL
<a href="#">CL60 (CL600)</a>	5731	9303	15667	21961	-	-	-	-	-	500 mL
<a href="#">CL68 (CL680)</a>	-	11093	18889		-	-	-	-	-	500 mL
<a href="#">CL74 (CL740)</a>	7585	12452	-		-	-	-	-	-	500 mL
<a href="#">VIS-CCS-14</a>	-	-	-		-	-	-	-	-	14 x 500 mL
<a href="#">VIS-CCS-18</a>	-	-	-		-	-	-	-	-	18 x 500 mL
<a href="#">VIS-CCS-8</a>	-	-	-		-	-	-	-	-	8 x 500 mL



## Section 1: Viscosity Reference Standards

Paragon Scientific – Viscosity Reference Standards – Cone & Plate		
Product Number	Product Description	Size (ml)
<a href="#">ALK-2162/1</a>	Cone & Plate (50 mPa·s @ 25 °C)	500ml
<a href="#">ALK-2162/2</a>	Cone & Plate (75 mPa·s @ 25 °C)	500ml
<a href="#">ALK-2162/3</a>	Cone & Plate (100 mPa·s @ 25 °C)	500ml
<a href="#">ALK-2162/4</a>	Cone & Plate (150 mPa·s @ 25 °C)	500ml
<a href="#">ALK-2162/5</a>	Cone & Plate (200 mPa·s @ 25 °C)	500ml
<a href="#">ALK-2162/6</a>	Cone & Plate (250 mPa·s @ 25 °C)	500ml
<a href="#">ALK-2162/7</a>	Cone & Plate (300 mPa·s @ 25 °C)	500ml
<a href="#">ALK-2162/8</a>	Cone & Plate (350 mPa·s @ 25 °C)	500ml
<a href="#">ALK-2162/9</a>	Cone & Plate (400 mPa·s @ 25 °C)	500ml
<a href="#">ALK-2162/10</a>	Cone & Plate (500 mPa·s @ 25 °C)	500ml
<a href="#">ALK-2162/11</a>	Cone & Plate (750 mPa·s @ 25 °C)	500ml
<a href="#">ALK-2162/12</a>	Cone & Plate (1000 mPa·s @ 25 °C)	500ml
<a href="#">ALK-2162/13</a>	Cone & Plate (1500 mPa·s @ 25 °C)	500ml
<a href="#">ALK-2162/14</a>	Cone & Plate (2500 mPa·s @ 25 °C)	500ml
<a href="#">ALK-2162/15</a>	Cone & Plate (3900 mPa·s @ 25 °C)	500ml
<a href="#">ALK-2162/16</a>	Cone & Plate (6000 mPa·s @ 25 °C)	500ml
<a href="#">ALK-2162/17</a>	Cone & Plate (7750 mPa·s @ 25 °C)	500ml
<a href="#">ALK-2162/18</a>	Cone & Plate (10000 mPa·s @ 25 °C)	500ml
<a href="#">ALK-2162/19</a>	Cone & Plate (15000 mPa·s @ 25 °C)	500ml
<a href="#">ALK-2162/20</a>	Cone & Plate (20000 mPa·s @ 25 °C)	500ml
<a href="#">ALK-2162/21</a>	Cone & Plate (14 mPa·s @ 25 °C)	500ml
<a href="#">ALK-2162/22</a>	Cone & Plate (28 mPa·s @ 25 °C)	500ml
<a href="#">ALK-2162/24</a>	Cone & Plate (2000 mPa·s @ 25 °C)	500ml

## Section 1: Viscosity Reference Standards

Cone and Plate Viscosity Reference Standards						
Part No.	Kinematic Viscosity mm <sup>2</sup> /s (cSt)					
	20.00 °C	21.00 °C	22.00 °C	23.00 °C	24.00 °C	25.00 °C
<a href="#">2162/1</a>	86.53	81.83	77.45	73.37	69.55	65.98
<a href="#">2162/2</a>	117.1	110.4	104.2	98.40	93.04	88.05
<a href="#">2162/3</a>	155.3	146.0	137.5	129.5	122.1	115.20
<a href="#">2162/4</a>	241.7	226.4	212.3	199.3	187.2	176.0
<a href="#">2162/5</a>	323.0	301.9	282.4	264.5	247.9	232.5
<a href="#">2162/6</a>	407.4	380.1	355.0	331.9	310.5	290.8
<a href="#">2162/7</a>	491.1	457.6	426.7	398.4	372.4	348.5
<a href="#">2162/8</a>	575.6	535.7	499.0	465.4	434.4	405.9
<a href="#">2162/9</a>	647.1	601.9	560.4	522.3	487.3	455.1
<a href="#">2162/10</a>	817.8	760.2	707.4	658.9	614.3	573.3
<a href="#">2162/11</a>	1240	1147	1063	985.9	915.3	850.7
<a href="#">2162/12</a>	1717	1587	1468	1360	1261	1170
<a href="#">2162/13</a>	2614	2409	2223	2054	1899	1757
<a href="#">2162/14</a>	4489	4127	3799	3501	3228	2980
<a href="#">2162/15</a>	7015	6435	5910	5433	4999	4605
<a href="#">2162/16</a>	10481	9595	8794	8069	7410	6813
<a href="#">2162/17</a>	14010	12812	11731	10752	9868	9066
<a href="#">2162/18</a>	17322	15835	14492	13278	12175	11176
<a href="#">2162/19</a>	26904	24576	22474	20574	18794	17187
<a href="#">2162/20</a>	37745	34358	31311	28567	26072	23822
<a href="#">2162/21</a>	20.92	20.06	19.24	18.47	17.74	17.05
<a href="#">2162/22</a>	42.73	40.67	38.75	36.94	35.24	33.65
<a href="#">2162/24</a>	3399	3129	2884	2661	2457	2271

## Section 1: Viscosity Reference Standards

Cone and Plate Viscosity Reference Standards						
Part No.	Dynamic Viscosity mPa·s (cP)					
	20.00 °C	21.00 °C	22.00 °C	23.00 °C	24.00 °C	25.00 °C
<a href="#">2162/1</a>	74.13	70.05	66.25	62.72	59.40	56.31
<a href="#">2162/2</a>	100.5	94.67	89.29	84.25	79.61	75.28
<a href="#">2162/3</a>	133.7	125.60	118.2	111.30	104.8	98.83
<a href="#">2162/4</a>	209.2	195.8	183.5	172.1	161.6	151.8
<a href="#">2162/5</a>	280.1	261.6	244.5	228.8	214.3	200.9
<a href="#">2162/6</a>	354.4	330.5	308.4	288.2	269.4	252.1
<a href="#">2162/7</a>	427.6	398.2	371.1	346.2	323.4	302.4
<a href="#">2162/8</a>	501.7	466.6	434.3	404.8	377.6	352.6
<a href="#">2162/9</a>	562.9	523.2	486.8	453.4	422.7	394.5
<a href="#">2162/10</a>	713.4	662.7	616.2	573.6	534.4	498.4
<a href="#">2162/11</a>	1083	1001	926.9	859.2	797.0	740.3
<a href="#">2162/12</a>	1503	1388	1283	1188	1101	1021
<a href="#">2162/13</a>	2294	2113	1948	1799	1662	1537
<a href="#">2162/14</a>	3951	3630	3340	3076	2834	2615
<a href="#">2162/15</a>	6190	5674	5208	4785	4400	4050
<a href="#">2162/16</a>	9274	8484	7771	7126	6540	6009
<a href="#">2162/17</a>	12417	11348	10384	9511	8724	8010
<a href="#">2162/18</a>	15368	14039	12841	11758	10775	9884
<a href="#">2162/19</a>	23912	21828	19950	18251	16663	15228
<a href="#">2162/20</a>	33604	30568	27842	25385	23155	21142
<a href="#">2162/21</a>	17.66	16.92	16.22	15.56	14.93	14.34
<a href="#">2162/22</a>	36.11	34.35	32.70	31.15	29.69	28.33
<a href="#">2162/24</a>	2989	2749	2532	2335	2155	1990

## Section 1: Viscosity Reference Standards

Cone and Plate Viscosity Reference Standards						
Part No.	Krebs (KU)					
	20.00 °C	21.00 °C	22.00 °C	23.00 °C	24.00 °C	25.00 °C
<a href="#">2162/1</a>	44.10	43.70	43.40	43.10	42.80	42.50
<a href="#">2162/2</a>	46.30	45.80	45.40	44.90	44.60	44.20
<a href="#">2162/3</a>	48.90	48.20	47.70	47.10	46.60	46.10
<a href="#">2162/4</a>	54.10	53.20	52.40	51.60	50.90	50.20
<a href="#">2162/5</a>	58.40	57.30	56.30	55.30	54.40	53.40
<a href="#">2162/6</a>	62.40	61.10	60.00	58.80	57.80	56.70
<a href="#">2162/7</a>	65.90	64.60	63.20	62.00	60.80	59.60
<a href="#">2162/8</a>	69.20	67.70	66.20	64.90	63.50	62.30
<a href="#">2162/9</a>	71.70	70.10	68.60	67.10	65.70	64.40
<a href="#">2162/10</a>	77.30	75.50	73.80	72.10	70.60	69.10
<a href="#">2162/11</a>	88.10	86.00	83.90	81.90	80.00	78.20
<a href="#">2162/12</a>	97.50	95.10	92.90	90.70	88.60	86.50
<a href="#">2162/13</a>	110.7	107.8	105.3	102.9	100.5	98.10
<a href="#">2162/14</a>	131.1	127.7	124.5	121.4	118.3	115.3
<a href="#">2162/15</a>	150.0	146.2	142.5	139.0	135.5	132.1
<a href="#">2162/16</a>	168.5	164.3	160.3	156.3	152.5	148.7
<a href="#">2162/17</a>	182.6	178.2	173.9	169.7	165.6	161.7
<a href="#">2162/18</a>	193.3	188.8	184.3	180.0	175.7	171.5
<a href="#">2162/19</a>	216.3	211.5	206.8	202.1	197.5	192.9
<a href="#">2162/20</a>	234.7	229.5	224.5	219.5	214.6	209.8
<a href="#">2162/21</a>	38.80	38.80	38.7	38.60	38.60	38.50
<a href="#">2162/22</a>	40.60	40.50	40.30	40.20	40.00	39.90
<a href="#">2162/24</a>	120.3	117.2	114.2	111.3	108.5	106.0

## Section 1: Viscosity Reference Standards

Cone and Plate Viscosity Reference Standards						
Part No.	Density g/mL					
	20.00 °C	21.00 °C	22.00 °C	23.00 °C	24.00 °C	25.00 °C
<a href="#">2162/1</a>	0.8567	0.8561	0.8554	0.8548	0.8541	0.8535
<a href="#">2162/2</a>	0.8581	0.8575	0.8569	0.8562	0.8556	0.8550
<a href="#">2162/3</a>	0.8610	0.8604	0.8598	0.8591	0.8585	0.8579
<a href="#">2162/4</a>	0.8655	0.8649	0.8643	0.8636	0.8630	0.8624
<a href="#">2162/5</a>	0.8671	0.8665	0.8659	0.8652	0.8646	0.8640
<a href="#">2162/6</a>	0.8700	0.8694	0.8688	0.8682	0.8676	0.8670
<a href="#">2162/7</a>	0.8708	0.8702	0.8696	0.8689	0.8683	0.8677
<a href="#">2162/8</a>	0.8716	0.8710	0.8704	0.8698	0.8692	0.8686
<a href="#">2162/9</a>	0.8699	0.8693	0.8687	0.8681	0.8675	0.8669
<a href="#">2162/10</a>	0.8723	0.8717	0.8711	0.8705	0.8699	0.8693
<a href="#">2162/11</a>	0.8732	0.8726	0.8720	0.8715	0.8708	0.8702
<a href="#">2162/12</a>	0.8755	0.8749	0.8743	0.8737	0.8731	0.8725
<a href="#">2162/13</a>	0.8777	0.8771	0.8765	0.8760	0.8754	0.8748
<a href="#">2162/14</a>	0.8802	0.8796	0.8791	0.8785	0.8780	0.8774
<a href="#">2162/15</a>	0.8824	0.8818	0.8812	0.8807	0.8801	0.8795
<a href="#">2162/16</a>	0.8848	0.8842	0.8837	0.8831	0.8826	0.8820
<a href="#">2162/17</a>	0.8863	0.8857	0.8852	0.8846	0.8841	0.8835
<a href="#">2162/18</a>	0.8872	0.8866	0.8861	0.8855	0.8850	0.8844
<a href="#">2162/19</a>	0.8888	0.8882	0.8877	0.8871	0.8866	0.8860
<a href="#">2162/20</a>	0.8903	0.8897	0.8892	0.8886	0.8881	0.8875
<a href="#">2162/21</a>	0.8443	0.8437	0.8430	0.8424	0.8417	0.8411
<a href="#">2162/22</a>	0.8451	0.8445	0.8438	0.8432	0.8425	0.8419
<a href="#">2162/24</a>	0.8793	0.8787	0.8781	0.8775	0.8769	0.8763

## Section 1: Viscosity Reference Standards

Paragon Scientific – Viscosity Reference Standards – Flow Cup		
Product Number	Product Description	Size (ml)
<a href="#">ALK-C200</a>	Viscosity Reference Standard Flow Cup Viscosity Standard type C200	500ml
<a href="#">ALK-C35</a>	Viscosity Reference Standard Flow Cup Viscosity Standard type C35	500ml
<a href="#">ALK-C350</a>	Viscosity Reference Standard Flow Cup Viscosity Standard type C350	500ml
<a href="#">ALK-C6</a>	Viscosity Reference Standard Flow Cup Viscosity Standard type C6	500ml
<a href="#">ALK-C60</a>	Viscosity Reference Standard Flow Cup Viscosity Standard type C60	500ml
<a href="#">ALK-C600</a>	Viscosity Reference Standard Flow Cup Viscosity Standard type C600	500ml

Flow Cup Viscosity Reference Standards 20 °C													
Part No.	Kinematic Viscosity mm <sup>2</sup> /s (cSt)	Dynamic Viscosity mPa·s (cP)	Density (g/mL)	DIN CUP		FORD CUP		ISO CUP		SHELL CUP		ZAHN CUP	
				Size (mm)	Drain Time(s)	Size (mm)	Drain Time(s)	Size (mm)	Drain Time(s)	Size (mm)	Drain Time(s)	Size (mm)	Drain Time(s)
<a href="#">ALK-C6</a>	10.46	8.703	0.832	-	-	1	56.3	3	36.1	1 2	59.3 23.2	1	38.5
<a href="#">ALK-C10</a>	20.92	17.66	0.8443	-	-	1	77.7	3	55.4	2 2.5	41.3 25.6	1	48
<a href="#">ALK-C20</a>	42.73	36.71	0.8451	-	-	2	47.7	4	35.3	2.5 3	49.2 30.3	1	67.8
<a href="#">ALK-C35</a>	86.53	74.13	0.8567	-	-	2 3	78.1 44.0	4	65.4	3 3.5	59.3 41.4	2	38.7
<a href="#">ALK-C60</a>	155.3	133.7	0.861	4	36.7	3 4	73.8 44.8	5	48.7	3.5 4	73.1 46.0	2 3	58.4 20.8
<a href="#">ALK-C100</a>	329.6	285.6	0.8664	4	73.5	4	90.1	5 6	101.2 49.4	5 6	51.7 20.8	3 4	35.7 27.3
<a href="#">ALK-C200</a>	647.1	562.9 -	0.8699 -	-	-	5	55.5	6	94.7	6	40.4	3 4	62.8 48.7
<a href="#">ALK-C350</a>	1240	1083	0.8732	-	-	-	-	-	-	6	77	5	53.9
<a href="#">ALK-C600</a>	2161	1894	0.8765	-	-	-	-	-	-	-	-	-	-

## Section 1: Viscosity Reference Standards

Flow Cup Viscosity Reference Standards 25 °C													
Part No.	Kinematic Viscosity mm <sup>2</sup> /s (cSt)	Dynamic Viscosity mPa·s (cP)	Density (g/mL)	DIN CUP		FORD CUP		ISO CUP		SHELL CUP		ZAHN CUP	
				Size (mm)	Drain Time(s)	Size (mm)	Drain Time(s)	Size (mm)	Drain Time(s)	Size (mm)	Drain Time(s)	Size (mm)	Drain Time(s)
<a href="#">ALK-C6</a>	8.869	7.350	0.8287	-	-	-	-	3	33.5	1 2	52.2 20.4	1	37.1
<a href="#">ALK-C10</a>	17.05	14.34	0.8411	-	-	1	69.8	3	47.9	2 2.5	34.6 21.4	1	44.5
<a href="#">ALK-C20</a>	33.65	28.33	0.8419	-	-	2	41.4	3	81.5	2.5 3	39.4 24.3	1 2	59.6 23.6
<a href="#">ALK-C35</a>	65.98	56.31	0.8535	-	-	2 3	63.8 35.1	4	51.0	2.5 3	74.3 45.7	2	32.9
<a href="#">ALK-C60</a>	115.2	98.83	0.8579	4	28.7	2 3	98.0 56.5	4 5	85.8 36.9	3.5 4	54.6 34.4	2	46.9
<a href="#">ALK-C100</a>	237.2	204.8	0.8634	4	53.7	4	66.1	5 6	73.2 36.6	4 5	69.8 37.5	3 4	27.8 21.0
<a href="#">ALK-C200</a>	455.1	394.5	0.8669	4	100.6	5	39.6	6	67.2	5 6	71.0 28.6	3 4	46.5 35.8
<a href="#">ALK-C350</a>	850.7	740.3	0.8702	-	-	5	72.3	-	-	6	53.0	4 5	62.5 37.0
<a href="#">ALK-C600</a>	1460	1275	0.8736	-	-	-	-	-	-	-	-	5	63.5



## Section 1: Viscosity Reference Standards

Paragon Scientific – Viscosity Reference Standards – High Temperature		
Product Number	Product Description	Size (ml)
<a href="#">ALK-N600</a>	Viscosity Reference Standard High Temperature Type N600	500ml
<a href="#">ALK-S2000S</a>	Viscosity Reference Standard High Temperature Type S2000S	500ml
<a href="#">ALK-S200HT</a>	Viscosity Reference Standard High Temperature Type S200HT, Temperature Ranges of 100 to 150 °C	500ml
<a href="#">ALK-S200S</a>	Viscosity Reference Standard High Temperature Type S200S	500ml
<a href="#">ALK-S20S</a>	Viscosity Reference Standard High Temperature Type S20S	500ml
<a href="#">ALK-S30000S</a>	Viscosity Reference Standard High Temperature Type N30000S	500ml
<a href="#">ALK-S3S</a>	Viscosity Reference Standard High Temperature Type S3S	500ml
<a href="#">ALK-S600HT</a>	Viscosity Reference Standard High Temperature Type S600HT, Temperature Ranges of 100 to 150 °C	500ml
<a href="#">ALK-S600S</a>	Viscosity Reference Standard High Temperature Type S600S	500ml
<a href="#">ALK-S60S</a>	Viscosity Reference Standard High Temperature Type S60S	500ml
<a href="#">ALK-S6S</a>	Viscosity Reference Standard High Temperature Type S6S	500ml
<a href="#">ALK-S8000S</a>	Viscosity Reference Standard High Temperature Type S8000S	500ml

High Temperature Viscosity Reference Standards												
Part No.	Kinematic Viscosity mm <sup>2</sup> /s (cSt)											
	°C	20.00	25.00	37.78	40.00	50.00	60.00	80.00	98.89	100.00	135.00	150.00
	°F	68.00	77.00	100.00	104.00	122.00	140.00	176.00	210.00	212.00	275.00	302.00
<a href="#">ALK-S3S</a>	4.601	4.058	3.055	2.920	2.417	2.041	1.522	1.207	1.192	-	-	0.7373
<a href="#">ALK-S6S</a>	10.46	8.869	6.117	5.772	4.528	3.655	2.540	1.913	1.885	-	-	1.073
<a href="#">ALK-S20S</a>	42.73	33.65	19.68	18.10	12.82	9.476	5.734	3.924	3.845	-	-	1.864
<a href="#">ALK-S60S</a>	155.3	115.2	58.80	52.93	34.24	23.41	12.47	7.773	7.582	-	-	3.155
<a href="#">ALK-N100S</a>	329.6	237.2	112.9	100.5	62.05	40.68	20.23	11.99	11.67	-	-	4.442
<a href="#">ALK-S200S</a>	647.1	455.1	205.1	180.9	107.5	68.05	31.88	18.08	17.54	-	-	6.178
<a href="#">ALK-S600S</a>	2161	1460	595.4	516.9	286.2	169.9	71.19	37.05	35.81	-	-	10.79
<a href="#">ALK-N600</a>	-	1460	-	-	-	169.9	-	-	-	14.59	-	-
<a href="#">ALK-S2000S</a>	8573	5566	2053	1752	899.4	497.2	182.9	85.90	82.53	-	-	20.51
<a href="#">ALK-N2000</a>	-	5566	-	-	-	497.2	-	-	-	29.11	-	-
<a href="#">ALK-S8000S</a>	37745	23822	8136	6853	3311	1724	560.5	235.9	225.3	-	-	44.56
<a href="#">ALK-S30000S</a>	138056	85267	27723	23124	10717	5334	1586	612.3	581.5	-	-	93.53



## Section 1: Viscosity Reference Standards

High Temperature Viscosity Reference Standards												
Part No.	Dynamic Viscosity mPa·s (cP)											
	°C	20.00	25.00	37.78	40.00	50.00	60.00	80.00	98.89	100.00	135.00	150.00
	°F	68.00	77.00	100.00	104.00	122.00	140.00	176.00	210.00	212.00	275.00	302.00
<a href="#">ALK-S3S</a>	3.780	3.320	2.472	2.358	1.936	1.620	1.187	0.9255	0.9131	-	0.5385	
<a href="#">ALK-S6S</a>	8.703	7.350	5.017	4.725	3.677	2.943	2.012	1.491	1.467	-	0.7988	
<a href="#">ALK-S20S</a>	36.11	28.33	16.41	15.07	10.59	7.767	4.627	3.120	3.054	-	1.421	
<a href="#">ALK-S60S</a>	133.7	98.83	49.97	44.92	28.84	19.58	10.27	6.313	6.153	-	2.463	
<a href="#">ALK-N100S</a>	285.6	204.8	96.60	85.84	52.62	34.25	16.79	9.813	9.544	-	3.498	
<a href="#">ALK-S200S</a>	562.9	394.5	176.2	155.2	91.59	57.57	26.59	14.88	14.42	-	4.895	
<a href="#">ALK-S600S</a>	1894	1275	515.7	447.1	245.8	144.9	59.91	30.77	29.72	-	8.642	
<a href="#">ALK-N600</a>	-	1275	-	-	-	144.9	-	-	-	11.81	-	
<a href="#">ALK-S2000S</a>	7571	4899	1792	1527	779.0	427.8	155.3	72.03	69.14	-	16.60	
<a href="#">ALK-N2000</a>	-	4899	-	-	-	427.8	-	-	-	23.81	-	
<a href="#">ALK-S8000S</a>	33604	21142	7164	6026	2893	1497	480.5	199.8	190.7	-	36.49	
<a href="#">ALK-S30000S</a>	123698	76169	24576	20469	9429	4665	1370	522.8	496.1	-	77.33	

High Temperature Viscosity Reference Standards						
Part no	Kinematic Viscosity mm <sup>2</sup> /s (cSt)					
	100 °C	110 °C	120 °C	130 °C	140 °C	150 °C
	212 °F	230 °F	248 °F	266 °F	284 °F	302 °F
<a href="#">ALKN100HT</a>	11.57	9.174	7.433	6.146	5.161	4.404
<a href="#">ALKS200HT</a>	17.54	13.64	10.87	8.850	7.349	6.178
<a href="#">ALKS600HT</a>	35.81	26.79	20.63	16.27	13.13	10.79

## Section 1: Viscosity Reference Standards

### High Temperature Viscosity Reference Standards

Part no	Dynamic Viscosity mPa.Ts (cP)					
	100 °C	110 °C	120 °C	130 °C	140 °C	150 °C
	212 °F	230 °F	248 °F	266 °F	284 °F	302 °F
<a href="#">ALKN100HT</a>	9.461	7.446	5.988	4.914	4.096	3.468
<a href="#">ALKS200HT</a>	14.42	11.13	8.808	7.118	5.867	4.895
<a href="#">ALKS600HT</a>	29.72	22.08	16.88	13.22	10.59	8.642

### High Temperature Viscosity Reference Standards

Part no	Density g/mL					
	100 °C	110 °C	120 °C	130 °C	140 °C	150 °C
	212 °F	230 °F	248 °F	266 °F	284 °F	302 °F
<a href="#">ALKN100HT</a>	0.81774	0.81165	0.80564	0.79956	0.79355	0.78742
<a href="#">ALKS200HT</a>	0.82216	0.81629	0.81027	0.80434	0.79836	0.79232
<a href="#">ALKS600HT</a>	0.82993	0.82422	0.81845	0.81262	0.80672	0.80092

### Paragon Scientific – Viscosity Reference Standards – Low Temperature

Product Number	Product Description	Size (ml)
<a href="#">ALK-N105B</a>	Viscosity Reference Standard Low Temperature Type N105B	500ml
<a href="#">ALK-N115B</a>	Viscosity Reference Standard Low Temperature Type N115B	500ml
<a href="#">ALK-N120B</a>	Viscosity Reference Standard Low Temperature Type N120B	500ml
<a href="#">ALK-N1400B</a>	Viscosity Reference Standard Low Temperature Type N1400B	500ml
<a href="#">ALK-N14B</a>	Viscosity Reference Standard Low Temperature Type N14B	500ml
<a href="#">ALK-N27B</a>	Viscosity Reference Standard Low Temperature Type N27B	500ml
<a href="#">ALK-N2B</a>	Viscosity Reference Standard Low Temperature Type N2B	500ml
<a href="#">ALK-N400B</a>	Viscosity Reference Standard Low Temperature Type N400B	500ml
<a href="#">ALK-N480B</a>	Viscosity Reference Standard Low Temperature Type N480B	500ml

## Section 1: Viscosity Reference Standards

Low Temperature Viscosity Reference Standards					
Part No.	Temperature		Kinematic Viscosity	Dynamic Viscosity	Density
	°C	°F	mm <sup>2</sup> /s (cSt)	mPa·s (cP)	g/mL
<a href="#">ALK-J10</a>	-40	-40	894.8	757	0.8460
<a href="#">ALK-N2B</a>	0	32	4.692	3.873	0.8254
	-20	-4	9.095	7.635	0.8395
<a href="#">ALK-N14B</a>	-25	-13	3393	2908	0.8570
	-30	-22	5761	4956	0.8602
	-35	-31	10300	8893	0.8634
	-40	-40	19605	16992	0.8667
<a href="#">ALK-N27B</a>	-17.78	0	2017	1722	0.8535
	-23.33	-10	3437	2946	0.8570
	-26.11	-15	4580	3933	0.8588
	-28.89	-20	6189	5326	0.8606
	-34.44	-30	11851	10242	0.8642
	-40	-40	24522	21278	0.8677
<a href="#">ALK-N105B</a>	-19	-2.2	30402	26398	0.8683
	-19.5	-3.1	32301	28057	0.8686
	-20	-4	34202	29718	0.8689
	-20.5	-4.9	36301	31553	0.8692
	-21	-5.8	38551	33520	0.8695
	-25	-13	63815	55647	0.8720
<a href="#">ALK-N115B</a>	-6.67	20	11155	9619	0.8623
	-12.22	10	19690	17044	0.8656
	-17.78	0	36685	31883	0.8691
	-23.33	-10	73165	63844	0.8726
	-26.11	-15	106422	93034	0.8742
	-28.89	-20	158599	138933	0.8760
<a href="#">ALK-N120B</a>	-39	-38.2	146812	128534	0.8755
	-39.5	-39.1	159232	139455	0.8758
	-40	-40	172601	151216	0.8761

Continued on the next page

## Section 1: Viscosity Reference Standards

Continued from previous page

Low Temperature Viscosity Reference Standards					
Part No.	Temperature		Kinematic Viscosity	Dynamic Viscosity	Density
	°C	°F	mm <sup>2</sup> /s (cSt)	mPa·s (cP)	g/mL
	-40.5	-40.9	187330	164176	0.8764
	-41	-41.8	203000	177970	0.8767
<a href="#">ALK-N400B</a>	-24	-11.2	59674	51982	0.8711
	-24.5	-12.1	63756	55563	0.8715
	-25	-13	68176	59436	0.8718
	-25.5	-13.9	72906	63581	0.8721
	-26	-14.8	77922	67979	0.8724
<a href="#">ALK-N480B</a>	-25	-13	148029	129511	0.8749
	-25.5	-13.9	158626	138845	0.8753
	-26	-14.8	170115	148936	0.8755
	-26.11	-15	172971	151471	0.8757
	-26.5	-15.7	182964	160240	0.8758
	-27	-16.6	196579	172243	0.8762
<a href="#">ALK-N1400B</a>	-11	12.2		133074	0.8837
	-11.5	11.3	150587	141732	0.8839
	-12	10.4	160349	150732	0.8841
	-12.22	10	170492	155081	0.8842
	-12.5	9.5	175391	160739	0.8843
	-13	8.6	181770	171074	0.8846
			193391		
<a href="#">ALK-JF1-H</a>	-20	-4	7.241	6.171	0.8522
	-40	-40	17.68	15.32	0.8664
<a href="#">ALK-JF1-L</a>	-20	-4	3.455	2.825	0.8178
	-40	-40	6.553	5.452	0.8320
ALK-P6000W-40C	-40	-40	6953	5989	0.8614
ALK-P130W-40C	-20	-4	36.33		
	-40	-40	132.3		

## Section 1: Viscosity Reference Standards

Paragon Scientific – Viscosity Reference Standards – Medical Grade		
Product Number	Product Description	Size (ml)
<a href="#">ALK-MGVS12-500</a>	Viscosity Reference Standard Medical grade viscosity standard, 1.2 mPa·s @ 25 °C - contains antibacterial agent	500ml
<a href="#">ALK-MGVS16-100</a>	Viscosity Reference Standard Medical grade viscosity standard, 1.6 mPa·s @ 25 °C - contains antibacterial agent	100ml
<a href="#">ALK-MGVS16-500</a>	Viscosity Reference Standard Medical grade viscosity standard, 1.6 mPa·s @ 25 °C - contains antibacterial agent	500ml
<a href="#">ALK-MGVS20-100</a>	Viscosity Reference Standard Medical grade viscosity standard, 2.0 mPa·s @ 25 °C - contains antibacterial agent	100ml
<a href="#">ALK-MGVS20-500</a>	Viscosity Reference Standard Medical grade viscosity standard, 2.0 mPa·s @ 25 °C - contains antibacterial agent	500ml
<a href="#">ALK-MGVS30-100</a>	Viscosity Reference Standard Medical grade viscosity standard, 3.0 mPa·s @ 25 °C - contains antibacterial agent	100ml
<a href="#">ALK-MGVS30-500</a>	Viscosity Reference Standard Medical grade viscosity standard, 3.0 mPa·s @ 25 °C - contains antibacterial agent	500ml
<a href="#">ALK-MGVS40-100</a>	Viscosity Reference Standard Medical grade viscosity standard, 4.0 mPa·s @ 25 °C - contains antibacterial agent	100ml
<a href="#">ALK-MGVS40-500</a>	Viscosity Reference Standard Medical grade viscosity standard, 4.0 mPa·s @ 25 °C - contains antibacterial agent	500ml
<a href="#">ALK-MGVS60-100</a>	Viscosity Reference Standard Medical grade viscosity standard, 6.0 mPa·s @ 25 °C - contains antibacterial agent	100ml
<a href="#">ALK-MGVS60-500</a>	Viscosity Reference Standard Medical grade viscosity standard, 6.0 mPa·s @ 25 °C - contains antibacterial agent	500ml



## Section 1: Viscosity Reference Standards

Medical Grade Viscosity Reference Standards			
Part No.	Dynamic Viscosity (mPa·s) 25.00 °C	Dynamic Viscosity (mPa·s) 37.00 °C	Pack Size
<a href="#">ALK-MGVS12-100</a>	1.201	0.9137	100 mL
<a href="#">ALK-MGVS12-500</a>	1.201	0.9137	500 mL
<a href="#">ALK-MGVS16-100</a>	1.606	1.197	100 mL
<a href="#">ALK-MGVS16-500</a>	1.606	1.197	500 mL
<a href="#">ALK-MGVS20-100</a>	2.002	1.467	100 mL
<a href="#">ALK-MGVS20-500</a>	2.002	1.467	500 mL
<a href="#">ALK-MGVS30-500</a>	3.002	2.131	500 mL
<a href="#">ALK-MGVS30-100</a>	3.002	2.131	100 mL
<a href="#">ALK-MGVS40-500</a>	4.009	2.773	500 mL
<a href="#">ALK-MGVS40-100</a>	4.009	2.773	100 mL
<a href="#">ALK-MGVS60-500</a>	6.027	4.016	500 mL
<a href="#">ALK-MGVS60-100</a>	6.027	4.016	100 mL
<a href="#">ALK-MGVS100-500</a>	9.994	6.335	500 mL
<a href="#">ALK-MGVS100-100</a>	9.994	6.335	100 mL



## Section 1: Viscosity Reference Standards

Paragon Scientific – Viscosity Reference Standards – Mineral Oil Rotational		
Product Number	Product Description	Size (ml)
<a href="#">ALK-RTM12</a>	Mineral Oil Rotational Viscosity Standard RTM12	500ml
<a href="#">ALK-RTM13</a>	Mineral Oil Rotational Viscosity Standard RTM13	500ml
<a href="#">ALK-RTM14</a>	Mineral Oil Rotational Viscosity Standard RTM14	500ml
<a href="#">ALK-RTM15</a>	Mineral Oil Rotational Viscosity Standard RTM15	500ml
<a href="#">ALK-RTM16</a>	Mineral Oil Rotational Viscosity Standard RTM16	500ml
<a href="#">ALK-RTM17</a>	Mineral Oil Rotational Viscosity Standard RTM17	500ml
<a href="#">ALK-RTM18</a>	Mineral Oil Rotational Viscosity Standard RTM18	500ml
<a href="#">ALK-RTM19</a>	Mineral Oil Rotational Viscosity Standard RTM19	500ml
<a href="#">ALK-RTM2</a>	Mineral Oil Rotational Viscosity Standard RTM2	500ml
<a href="#">ALK-RTM20</a>	Mineral Oil Rotational Viscosity Standard RTM20	500ml
<a href="#">ALK-RTM21</a>	Mineral Oil Rotational Viscosity Standard RTM21	500ml
<a href="#">ALK-RTM22</a>	Mineral Oil Rotational Viscosity Standard RTM22	500ml
<a href="#">ALK-RTM23</a>	Mineral Oil Rotational Viscosity Standard RTM23	500ml
<a href="#">ALK-RTM24</a>	Mineral Oil Rotational Viscosity Standard RTM24	500ml
<a href="#">ALK-RTM25</a>	Mineral Oil Rotational Viscosity Standard RTM25	500ml
<a href="#">ALK-RTM26</a>	Mineral Oil Rotational Viscosity Standard RTM26	500ml
<a href="#">ALK-RTM27</a>	Mineral Oil Rotational Viscosity Standard RTM27	500ml
<a href="#">ALK-RTM28</a>	Mineral Oil Rotational Viscosity Standard RTM28	500ml
<a href="#">ALK-RTM29</a>	Mineral Oil Rotational Viscosity Standard RTM29	500ml
<a href="#">ALK-RTM3</a>	Mineral Oil Rotational Viscosity Standard RTM3	500ml
<a href="#">ALK-RTM30</a>	Mineral Oil Rotational Viscosity Standard RTM30	500ml
<a href="#">ALK-RTM31</a>	Mineral Oil Rotational Viscosity Standard RTM30	500ml
<a href="#">ALK-RTM32</a>	Mineral Oil Rotational Viscosity Standard RTM32	500ml
<a href="#">ALK-RTM33</a>	Mineral Oil Rotational Viscosity Standard RTM33	500ml
<a href="#">ALK-RTM34</a>	Mineral Oil Rotational Viscosity Standard RTM34	500ml
<a href="#">ALK-RTM35</a>	Mineral Oil Rotational Viscosity Standard RTM35	500ml
<a href="#">ALK-RTM36</a>	Mineral Oil Rotational Viscosity Standard RTM36	500ml

Continued on the next page

## Section 1: Viscosity Reference Standards

Continued from previous page

Paragon Scientific – Viscosity Reference Standards – Mineral Oil Rotational		
Product Number	Product Description	Size (ml)
<a href="#">ALK-RTM37</a>	Mineral Oil Rotational Viscosity Standard RTM37	500ml
<a href="#">ALK-RTM38</a>	Mineral Oil Rotational Viscosity Standard RTM38	500ml
<a href="#">ALK-RTM39</a>	Mineral Oil Rotational Viscosity Standard RTM39	500ml
<a href="#">ALK-RTM4</a>	Mineral Oil Rotational Viscosity Standard RTM4	500ml
<a href="#">ALK-RTM5</a>	Mineral Oil Rotational Viscosity Standard RTM5	500ml
<a href="#">ALK-RTM6</a>	Mineral Oil Rotational Viscosity Standard RTM6	500ml
<a href="#">ALK-RTM7</a>	Mineral Oil Rotational Viscosity Standard RTM7	500ml
<a href="#">ALK-RTM8</a>	Mineral Oil Rotational Viscosity Standard RTM8	500ml
<a href="#">ALK-RTM9</a>	Mineral Oil Rotational Viscosity Standard RTM9	500ml

Mineral Oil Rotational Viscosity Reference Standards												
Part No.	Dynamic Viscosity mPa·s (cP)											
	°C	20.00	20.50	21.00	21.50	22.00	22.50	23.00	23.50	24.00	24.50	25.00
	°F	68.00	68.90	69.80	70.70	71.60	72.50	73.40	74.30	75.20	76.10	77.00
<a href="#">ALK-RTM1</a>	0.3094	0.3078	0.3063	0.3048	0.3033	0.3017	0.3002	0.2988	0.2973	0.2958	0.2944	
<a href="#">ALK-RTM2</a>	0.5188	0.5157	0.5126	0.5095	0.5064	0.5035	0.5004	0.4975	0.4945	0.4916	0.4887	
<a href="#">ALK-RTM3</a>	1.014	1.006	0.998	0.9907	0.9832	0.9757	0.9682	0.9615	0.9541	0.9466	0.9400	
<a href="#">ALK-RTM4</a>	2.427	2.400	2.373	2.347	2.321	2.296	2.271	2.247	2.223	2.199	2.176	
<a href="#">ALK-RTM5</a>	3.780	3.730	3.681	3.633	3.585	3.539	3.494	3.449	3.405	3.362	3.320	
<a href="#">ALK-RTM6</a>	5.446	5.364	5.283	5.204	5.126	5.051	4.976	4.903	4.832	4.762	4.695	
<a href="#">ALK-RTM7</a>	8.703	8.550	8.404	8.262	8.123	7.986	7.853	7.723	7.596	7.471	7.350	
<a href="#">ALK-RTM8</a>	12.17	11.94	11.71	11.49	11.27	11.07	10.86	10.66	10.46	10.28	10.09	
<a href="#">ALK-RTM9</a>	17.66	17.29	16.92	16.56	16.22	15.88	15.56	15.24	14.93	14.63	14.34	
<a href="#">ALK-RTM10</a>	25.69	25.10	24.52	23.96	23.42	22.89	22.39	21.90	21.42	20.95	20.50	
<a href="#">ALK-RTM11</a>	36.11	35.22	34.35	33.51	32.70	31.91	31.15	30.41	29.70	29.01	28.33	
<a href="#">ALK-RTM12</a>	74.13	72.06	70.05	68.11	66.24	64.44	62.71	61.03	59.39	57.83	56.31	
<a href="#">ALK-RTM13</a>	100.5	97.53	94.67	91.98	89.29	86.77	84.28	81.91	79.62	77.41	75.28	

Continued on the next page



## Section 1: Viscosity Reference Standards

Continued from previous page

Mineral Oil Rotational Viscosity Reference Standards												
Part No.	Dynamic Viscosity mPa·s (cP)											
	°C	20.00	20.50	21.00	21.50	22.00	22.50	23.00	23.50	24.00	24.50	25.00
	°F	68.00	68.90	69.80	70.70	71.60	72.50	73.40	74.30	75.20	76.10	77.00
<a href="#">ALK-RTM14</a>	133.7	129.6	125.6	121.8	118.1	114.7	111.3	108	104.8	101.8	98.83	
<a href="#">ALK-RTM15</a>	209.2	202.4	195.8	189.5	183.5	177.6	172	166.7	161.6	156.6	151.8	
<a href="#">ALK-RTM16</a>	281.8	272.4	263.3	254.5	246.2	238.2	230.4	223	215.9	209	202.5	
<a href="#">ALK-RTM17</a>	354.4	342.1	330.5	319.1	308.3	298	288.1	278.5	269.4	260.5	252.1	
<a href="#">ALK-RTM18</a>	427.6	412.6	398.2	384.4	371.1	358.5	346.3	334.7	323.4	312.7	302.4	
<a href="#">ALK-RTM19</a>	501.7	483.7	466.6	450.1	434.3	419.2	404.7	390.8	377.6	364.9	352.6	
<a href="#">ALK-RTM20</a>	562.9	542.6	523.2	504.6	486.8	469.8	453.4	437.7	422.7	408.4	394.5	
<a href="#">ALK-RTM21</a>	727.1	700.3	674.7	650.2	626.6	604.3	582.7	562.1	542.4	523.5	505.4	
<a href="#">ALK-RTM22</a>	1083	1041	1002	963	926.9	892.6	859.1	827.5	797	768	740.3	
<a href="#">ALK-RTM23</a>	1136	1098	1061	1027	993.7	961.2	930.4	900.5	871.4	844	817.4	
<a href="#">ALK-RTM24</a>	1503	1444	1388	1335	1283	1235	1188	1143	1101	1060	1021	
<a href="#">ALK-RTM25</a>	1894	1818	1747	1678	1611	1549	1489	1432	1378	1326	1275	
<a href="#">ALK-RTM26</a>	2294	2201	2113	2029	1948	1872	1798	1729	1662	1598	1537	
<a href="#">ALK-RTM27</a>	2989	2866	2749	2638	2532	2431	2334	2242	2154	2071	1990	
<a href="#">ALK-RTM28</a>	3951	3786	3630	3481	3339	3203	3074	2951	2833	2722	2615	
<a href="#">ALK-RTM29</a>	6190	5925	5674	5434	5206	4990	4783	4587	4399	4220	4050	
<a href="#">ALK-RTM30</a>	7571	7239	6923	6624	6340	6069	5811	5566	5333	5110	4899	
<a href="#">ALK-RTM32</a>	9274	8868	8483	8119	7770	7439	7124	6824	6540	6268	6009	
<a href="#">ALK-RTM33</a>	12119	11578	11066	10579	10116	9677	9260	8862	8484	8124	7782	
<a href="#">ALK-RTM34</a>	12417	11869	11348	10855	10385	9938	9513	9109	8725	8359	8010	
<a href="#">ALK-RTM35</a>	15368	14686	14038	13424	12839	12283	11753	11251	10773	10317	9884	
<a href="#">ALK-RTM36</a>	23912	22828	21799	20824	19897	19016	18179	17384	16630	15912	15228	
<a href="#">ALK-RTM37</a>	33604	32041	30559	29157	27825	26560	25361	24222	23144	22118	21142	
<a href="#">ALK-RTM38</a>	66571	63396	60396	57550	54852	52302	49879	47582	45403	43340	41379	
<a href="#">ALK-RTM39</a>	123698	117689	112016	106635	101543	96731	92163	87835	83735	79856	76169	

## Section 1: Viscosity Reference Standards

Paragon Scientific – Viscosity Reference Standards – Rotational Type		
Product Number	Product Description	Size (ml)
<a href="#">ALK-VIS-RT10K-600</a>	Viscosity Reference Standard Rotational Type RT10,000	600ml
<a href="#">ALK-VIS-RT12K-600</a>	Viscosity Reference Standard Rotational Type RT12500	600ml
<a href="#">ALK-VIS-RT1K-600</a>	Viscosity Reference Standard Rotational Type RT1000	600ml
<a href="#">ALK-VIS-RT250-600</a>	Viscosity Reference Standard Rotational Type RT250	600ml
<a href="#">ALK-VIS-RT30K-600</a>	Viscosity Reference Standard Rotational Type RT30000	600ml
<a href="#">ALK-VIS-RT350-600</a>	Viscosity Reference Standard Rotational Type RT350	600ml
<a href="#">ALK-VIS-RT5-600</a>	Viscosity Reference Standard Rotational Type RT5	600ml
<a href="#">ALK-VIS-RT50-600</a>	Viscosity Reference Standard Rotational Type RT50	600ml
<a href="#">ALK-VIS-RT500-600</a>	Viscosity Reference Standard Rotational Type RT500	600ml
<a href="#">ALK-VIS-RT5K-600</a>	Viscosity Reference Standard Rotational Type RT5000	600ml
<a href="#">ALK-VIS-RT60K-600</a>	Viscosity Reference Standard Rotational Type RT60000	600ml
<a href="#">ALK-VIS-RT75-600</a>	Viscosity Reference Standard Rotational Type RT75	600ml



## Section 1: Viscosity Reference Standards

Silicon Rotational Viscosity Reference Standards						
Part No.	Kinematic Viscosity mm <sup>2</sup> /s (cSt)		Dynamic Viscosity mPa·s (cP)		Density g/mL	
	20.00 °C	25.00 °C	20.00 °C	25.00 °C	20.00 °C	25.00 °C
	68.00 °F	77.00 °F	68.00 °F	77.00 °F	68.00 °F	77.00 °F
<a href="#">ALK-VIS-RT5-600</a>	5.743	5.278	5.278	4.826	0.9190	0.9143
<a href="#">ALK-VIS-RT10-600</a>	11.81	10.79	11.09	10.08	0.9389	0.9343
<a href="#">ALK-VIS-RT50-600</a>	58.53	53.42	55.14	50.08	0.9421	0.9375
<a href="#">ALK-VIS-RT75-600</a>	88.21	80.38	83.45	75.67	0.9460	0.9414
<a href="#">ALK-VIS-RT100-600</a>	115.2	104.3	111.5	100.5	0.9678	0.9633
<a href="#">ALK-VIS-RT250-600</a>	289.8	263.9	275.2	249.4	0.9495	0.9449
<a href="#">ALK-VIS-RT350-600</a>	410.4	373.4	390.1	353.2	0.9505	0.9459
<a href="#">ALK-VIS-RT500-600</a>	579.6	527	552.6	500	0.9534	0.9488
<a href="#">ALK-VIS-RT1K-600</a>	1157	1051	1107	1001	0.9567	0.9521
<a href="#">ALK-VIS-RT5K-600</a>	5840	5296	5651	5101	0.9677	0.9632
<a href="#">ALK-VIS-RT10K-600</a>	11341	10273	11034	9949	0.9729	0.9685
<a href="#">ALK-VIS-RT12K-600</a>	13745	12446	13399	12076	0.9748	0.9703
<a href="#">ALK-VIS-RT30K-600</a>	33389	30244	32541	29343	0.9746	0.9702
<a href="#">ALK-VIS-RT60K-600</a>	64845	58742	63217	57003	0.9749	0.9704
<a href="#">ALK-VIS-RT100K-600</a>	108039	97858	105327	94961	0.9749	0.9704

## Section 1: Viscosity Reference Standards

Paragon Scientific – Viscosity Reference Standards – Bath Media		
Product Number	Product Description	Size (l)
<a href="#">ALK-BM2-5L</a>	Viscosity Bath Media, White Oil for use 80°C to 120°C	5l
<a href="#">ALK-BM5-20L</a>	Viscosity Bath Media Silicone Fluid 26 cSt @ 25°C for use 120 to 150°C	20l
<a href="#">ALK-BM5-5L</a>	Viscosity Bath Media Silicone Fluid 26 cSt @ 25°C for use 120 to 150°C	5l
<a href="#">ALK-BM6-20L</a>	Viscosity Bath Media Silicone Fluid 20 cSt @ 25°C for use 90 to 135°C	20l
<a href="#">ALK-BM6-5L</a>	Viscosity Bath Media Silicone Fluid 20 cSt @ 25°C for use 90 to 135°C	5l
<a href="#">ALK-BM7-20L</a>	Viscosity Bath Media Silicone Fluid 10 cSt @ 25°C for use 50 to 100°C	20l
<a href="#">ALK-BM7-5L</a>	Viscosity Bath Media Silicone Fluid 10 cSt @ 25°C for use 50 to 100°C	5l
<a href="#">ALK-BM8-20L</a>	Viscosity Bath Media Silicone Fluid 5 cSt @ 25°C for use 20 to 60°C	20l
<a href="#">ALK-BM8-5L</a>	Viscosity Bath Media Silicone Fluid 5 cSt @ 25°C for use 20 to 60°C	5l

Paragon Scientific – Viscosity Reference Standards – Check Oil		
Product Number	Product Description	Size (ml)
<a href="#">ALK-CVCO15W40-5L</a>	Viscosity Check Oil, 15W40, Kinematic Viscosity, CCS Dynamic Viscosity & Density	5l
<a href="#">ALK-CVCO5W30</a>	Viscosity Check Oil, 5W30, Kinematic Viscosity, CCS Dynamic Viscosity & Density	500ml
<a href="#">ALK-CVCO5W30-5L</a>	Viscosity Check Oil, 5W30, Kinematic Viscosity, CCS Dynamic Viscosity & Density	5l

## Section 1: Viscosity Reference Standards

Paragon Scientific – Viscosity Reference Standards – Small Sample		
Product Number	Product Description	Size (ml)
<a href="#">ALK-HVS04</a>	Small Sample Viscosity Reference Standard, (cSt) @ 40 °C & 100 °C (32.42 cSt @ 40 °C)	60ml
<a href="#">ALK-HVS05</a>	Small Sample Viscosity Reference Standard, (cSt) @ 40 °C & 100 °C (54.29 cSt @ 40 °C)	60ml
<a href="#">ALK-HVS06</a>	Small Sample Viscosity Reference Standard, (cSt) @ 40 °C & 100 °C (100.2 cSt @ 40 °C)	60ml
<a href="#">ALK-HVS07</a>	Small Sample Viscosity Reference Standard, (cSt) @ 40 °C & 100 °C (183.0 cSt @ 40 °C)	60ml
<a href="#">ALK-HVS08</a>	Small Sample Viscosity Reference Standard, (cSt) @ 40 °C & 100 °C (306.9 cSt @ 40 °C)	60ml
<a href="#">ALK-HVS09</a>	Small Sample Viscosity Reference Standard, (cSt) @ 40 °C & 100 °C (528.1 cSt @ 40 °C)	60ml
<a href="#">ALK-HVS10</a>	Small Sample Viscosity Reference Standard, (cSt) @ 40 °C & 100 °C (1003 cSt @ 40 °C)	60ml
<a href="#">ALK-HVS11</a>	Small Sample Viscosity Reference Standard, (cSt) @ 40 °C & 100 °C (1706 cSt @ 40 °C)	60ml
<a href="#">ALK-HVS12</a>	Small Sample Viscosity Reference Standard, (cSt) @ 40 °C & 100 °C (2100 cSt @ 40 °C)	60ml
<a href="#">ALK-HVS13</a>	Small Sample Viscosity Reference Standard, (cSt) @ 40 °C & 100 °C (3420 cSt @ 40 °C)	60ml
<a href="#">ALK-HVS14</a>	Small Sample Viscosity Reference Standard, (cSt) @ 40 °C & 100 °C (6846 cSt @ 40 °C)	60ml
<a href="#">ALK-HVS15</a>	Small Sample Viscosity Reference Standard, (cSt) @ 40 °C & 100 °C (13014 cSt @ 40 °C)	60ml
<a href="#">ALK-HVS16</a>	Small Sample Viscosity Reference Standard, (cSt) @ 40 °C & 100 °C (23192 cSt @ 40 °C)	60ml

## Section 1: Viscosity Reference Standards

Small Sample Viscosity Reference Standards					
Part No.	Kinematic mm <sup>2</sup> /s (cSt) 40°C	Kinematic mm <sup>2</sup> /s (cSt) 100°C	Test Method	Accreditation	Pack Size
<a href="#">ALK-HVS01</a>	2.920	1.192	ASTM D2162	ISO 17025 / ISO 17034	60 mL
<a href="#">ALK-HVS02</a>	5.772	1.885	ASTM D2162	ISO 17025 / ISO 17034	60 mL
<a href="#">ALK-HVS03</a>	10.05	2.646	ASTM D2162	ISO 17025 / ISO 17034	60 mL
<a href="#">ALK-HVS04</a>	32.62	5.620	ASTM D2162	ISO 17025 / ISO 17034	60 mL
<a href="#">ALK-HVS05</a>	52.93	7.582	ASTM D2162	ISO 17025 / ISO 17034	60 mL
<a href="#">ALK-HVS06</a>	99.66	11.63	ASTM D2162	ISO 17025 / ISO 17034	60 mL
<a href="#">ALK-HVS07</a>	180.9	17.54	ASTM D2162	ISO 17025 / ISO 17034	60 mL
<a href="#">ALK-HVS08</a>	316.1	25.42	ASTM D2162	ISO 17025 / ISO 17034	60 mL
<a href="#">ALK-HVS09</a>	516.9	35.81	ASTM D2162	ISO 17025 / ISO 17034	60 mL
<a href="#">ALK-HVS10</a>	1001	57.51	ASTM D2162	ISO 17025 / ISO 17034	60 mL
<a href="#">ALK-HVS11</a>	1752	82.53	ASTM D2162	ISO 17025 / ISO 17034	60 mL
<a href="#">ALK-HVS12</a>	2100	95.45	ASTM D2162	ISO 17025 / ISO 17034	60 mL
<a href="#">ALK-HVS13</a>	3435	140.4	ASTM D2162	ISO 17025 / ISO 17034	60 mL
<a href="#">ALK-HVS14</a>	6853	225.3	ASTM D2162	ISO 17025 / ISO 17034	60 mL
<a href="#">ALK-HVS15</a>	12954	367.7	ASTM D2162	ISO 17025 / ISO 17034	60 mL
<a href="#">ALK-HVS16</a>	23124	581.5	ASTM D2162	ISO 17025 / ISO 17034	60 mL



## Section 1: Viscosity Reference Standards

VHG – Viscosity Reference Standards		
Product Number	Product Description	Size (ml)
<a href="#">VHG-VISC10-500</a>	Viscosity Reference Standard – 10 cSt at 40 °C, and 2.7 cSt at 100 °C	500ml
<a href="#">VHG-VISC100A-100</a>	Viscosity Reference Standard – approximately 100 cSt at 40 °C and approximately 16.8 cSt at 100 °C.	100ml
<a href="#">VHG-VISC100A-500</a>	Viscosity Reference Standard – 100 cSt at 40 °C, and 16.8 cSt at 100 °C	500ml
<a href="#">VHG-VISC120-500</a>	Viscosity Reference Standard – 120 cSt at 40 °C, and 20 cSt at 100 °C	500ml
<a href="#">VHG-VISC180-500</a>	Viscosity Reference Standard – 180 cSt at 40 °C, and 26 cSt at 100 °C	500ml
<a href="#">VHG-VISC20-500</a>	Viscosity Reference Standard – 19 cSt at 40 °C, and 5 cSt at 100 °C	500ml
<a href="#">VHG-VISC30-500</a>	Viscosity Reference Standard – 30 cSt at 40 °C, and 5.3 cSt at 100 °C	500ml
<a href="#">VHG-VISC360-500</a>	Viscosity Reference Standard – 360 cSt at 40 °C, and 42 cSt at 100 °C	500ml
<a href="#">VHG-VISC50-500</a>	Viscosity Reference Standard – 50 cSt at 40 °C, and 7.3 cSt at 100 °C	500ml
<a href="#">VHG-VISC500-50</a>	Viscosity Reference Standard – approximately 500 cSt at 40 °C and approximately 52 cSt at 100 °C.	50ml
<a href="#">VHG-VISC500-500</a>	Viscosity Reference Standard -500 cSt at 40 °C, and 52 cSt at 100 °C	500ml
<a href="#">VHG-VISC60A-500</a>	Viscosity Reference Standard -60 cSt at 40°C, and 11.4 cSt at 100°C	500ml
<a href="#">VHG-VISC75-500</a>	Viscosity Reference Standard -73 cSt at 40°C, and 9 cSt at 100°C	500ml
<a href="#">VHG-VISC900-500</a>	Viscosity Reference Standard -930 cSt at 40°C, and 82 cSt at 100°C	500ml



## Section 1: Viscosity Reference Standards

Pure Water Viscosity Reference Standards				
Part no	Kinematic Viscosity mm <sup>2</sup> /s (cSt)			
	5.00 °C	20.00 °C	25.00 °C	37.00 °C
	41.00 °F	68.00 °F	77.00 °F	98.60 °F
<a href="#">ALK-VISC-WAT</a>	1.5265	1.0035	0.8928	0.69599

Pure Water Viscosity Reference Standards				
Part no	Dynamic Viscosity mPa. s (cP)			
	5.00 °C	20.00 °C	25.00 °C	37.00 °C
	41.00 °F	68.00 °F	77.00 °F	98.60 °F
<a href="#">ALK-VISC-WAT</a>	1.5264	1.0017	0.89018	0.69135

Pure Water Viscosity Reference Standards				
Part no	Density g/mL			
	5.00 °C	20.00 °C	25.00 °C	37.00 °C
	41.00 °F	68.00 °F	77.00 °F	98.60 °F
<a href="#">ALK-VISC-WAT</a>	0.99996	0.99822	0.99706	0.99334



## Section 2

# Base Number Reference Standards

LGC Industrial's ISO 17025 & ISO 17034 total Base Number (TBN) Standards are specifically manufactured for the verification of analytical instruments used to determine base number by potentiometric titration. TBN standards are certified in accordance with ASTM D2896 / IP 276 and are applicable for use with other internationally equivalent methods.

The TBN value is important in the lubricants industry. Like the TAN Value, it is used as a confirmatory in the quality control of new and used lubricants, such as condition monitoring during lubricant use. TBN is derived from additives such as detergents and is an indicator of the lubricants ability to neutralise acids that can be formed during operation.

Paragon Scientific – Base Number Reference Standards – UKAS ISO 17025 / ISO 17034 Certified		
Product Number	Product Description	Size (g)
<a href="#">ALK-TBN1</a>	Base Number Standard Certified Value 1 mg KOH/g	125g
<a href="#">ALK-TBN1/3</a>	Base Number Standard Certified Value 1 mg KOH/g	3x125g
<a href="#">ALK-TBN10</a>	Base Number Standard Certified Value 15 mg KOH/g	50g
<a href="#">ALK-TBN10/3</a>	Base Number Standard Certified Value 15 mg KOH/g,	3x50g
<a href="#">ALK-TBN15</a>	Base Number Standard Certified Value 30 mg KOH/g	50g
<a href="#">ALK-TBN15/3</a>	Base Number Standard Certified Value 30 mg KOH/g	3x50g
<a href="#">ALK-TBN3</a>	Base Number Standard Certified Value 3 mg KOH/g	50g
<a href="#">ALK-TBN3/3</a>	Base Number Standard Certified Value 3mg KOH/g	3x50g
<a href="#">ALK-TBN30</a>	Base Number Standard Certified Value 40 mg KOH/g	50g
<a href="#">ALK-TBN30/3</a>	Base Number Standard Certified Value 40 mg KOH/g	3x50g
<a href="#">ALK-TBN40</a>	Base Number Standard Certified Value 6.0 mg KOH/g	50g
<a href="#">ALK-TBN40/3</a>	Base Number Standard Certified Value 6.0 mg KOH/g	3x50g
<a href="#">ALK-TBN6</a>	Base Number Standard Certified Value 10 mg KOH/g	50g
<a href="#">ALK-TBN6/3</a>	Base Number Standard Certified Value 10 mg KOH/g	3x50g
<a href="#">ALK-TBN70</a>	Base Number Standard Certified Value 70 mg KOH/g	50g
<a href="#">ALK-TBN70/3</a>	Base Number Standard Certified Value 70 mg KOH/g	3x50g

## Section 2: Base Number Reference Standards

VHG – Base Number Reference Standards		
Product Number	Product Description	Size (g)
<a href="#">VHG-BN-10-100G</a>	Base Number Standard 10 mgKOH/g	100g
<a href="#">VHG-BN-10-400G</a>	Base Number Standard 10 mgKOH/g	400g
<a href="#">VHG-BN-10-50G</a>	Base Number (BN) Standard: 10 mg KOH/g in Hydrocarbon Oil	50g
<a href="#">VHG-BN-15-50G</a>	Base Number (BN) Standard: 15 mg KOH/g in Hydrocarbon Oil	50g
<a href="#">VHG-BN-30-100G</a>	Base Number Standard 30 mgKOH/g	100g
<a href="#">VHG-BN-30-400G</a>	Base Number Standard 30 mgKOH/g	400g
<a href="#">VHG-BN-30-50G</a>	Base Number (BN) Standard: 30 mg KOH/g in Hydrocarbon Oil	50g
<a href="#">VHG-BN-40-100G</a>	Base Number (BN) Standard: 40 mg KOH/g in Hydrocarbon Oil	100g
<a href="#">VHG-BN-40-50G</a>	Base Number (BN) Standard: 40 mg KOH/g in Hydrocarbon Oil	50g
<a href="#">VHG-BN-6-100G</a>	Base Number Standard 6 mgKOH/g	100g
<a href="#">VHG-BN-6-400G</a>	Base Number Standard 6 mgKOH/g	400g
<a href="#">VHG-BN-6-50G</a>	Base Number (BN) Standard: 6 mg KOH/g in Hydrocarbon Oil	50g
<a href="#">VHG-BN-6-800G</a>	Base Number Standard 6 mgKOH/g	800g
<a href="#">VHG-BN-70-50G</a>	Base Number (BN) Standard: 70 mg KOH/g in Hydrocarbon Oil	50g



## Section 3

# Acid Number Reference Standards

LGC Industrial's ISO 17025 & ISO 17034 Total Acid Number (TAN) Standards manufactured for the verification of analytical instruments used to determine acid number by potentiometric titration. TAN Standards are certified in accordance with ASTM D664 / IP 177 and are applicable for use with internationally equivalent methods.

The TAN value is important in the lubricants industry as a confirmatory test in the quality control testing of new lubricants and is also used for condition monitoring of in use / used lubricants. The total acid number will be an indicator for the age of a lubricant and assist on the timing of oil changes.

Paragon Scientific – Acid Number Reference Standards – UKAS ISO 17025 / ISO 17034 Certified		
Product Number	Product Description	Size (g)
<a href="#">ALK-TAN001</a>	Acid Number Standard Certified Value 0.1 mg KOH/g	125g
<a href="#">ALK-TAN001/3</a>	Acid Number Standard Certified Value 0.1 mg KOH/g	3x125g
<a href="#">ALK-TAN005</a>	Acid Number Standard Certified Value 0.5 mg KOH/g	125g
<a href="#">ALK-TAN005/3</a>	Acid Number Standard Certified Value 0.5 mg KOH/g	3x125g
<a href="#">ALK-TAN010</a>	Acid Number Standard Certified Value 1.0 mg KOH/g	125g
<a href="#">ALK-TAN010/3</a>	Acid Number Standard Certified Value 1.0 mg KOH/g	3x125g
<a href="#">ALK-TAN015</a>	Acid Number Standard Certified Value 1.5 mg KOH/g	125g
<a href="#">ALK-TAN015/3</a>	Acid Number Standard Certified Value 1.5 mg KOH/g	3x125g
<a href="#">ALK-TAN020</a>	Acid Number Standard Certified Value 2.0 mg KOH/g	50g
<a href="#">ALK-TAN020/3</a>	Acid Number Standard Certified Value 2.0 mg KOH/g	3x50g
<a href="#">ALK-TAN025</a>	Acid Number Standard Certified Value 2.5 mg KOH/g	50g
<a href="#">ALK-TAN025/3</a>	Acid Number Standard Certified Value 2.5 mg KOH/g	3x50g
<a href="#">ALK-TAN030</a>	Acid Number Standard Certified Value 3.0 mg KOH/g	50g
<a href="#">ALK-TAN030/3</a>	Acid Number Standard Certified Value 3.0 mg KOH/g	3x50g
<a href="#">ALK-TAN050</a>	Acid Number Standard Certified Value 4.57 mg KOH/g	50g
<a href="#">ALK-TAN050/3</a>	Acid Number Standard Certified Value 4.57 mg KOH/g	3x50g
<a href="#">ALK-TANI00</a>	Acid Number Standard Certified Value 10.14 mg KOH/g	50g
<a href="#">ALK-TANI00/3</a>	Acid Number Standard Certified Value 10.14 mg KOH/g	3x50g

### Section 3: Acid Number Reference Standards

VHG – Acid Number Reference Standards		
Product Number	Product Description	Size (g)
<a href="#">VHG-AN-0.1-100G</a>	Acid Number (AN) Standard: 0.1 mg KOH/g in Hydrocarbon Oil	100g
<a href="#">VHG-AN-0.1-400G</a>	Acid Number Standard 0.1 mgKOH/g	400g
<a href="#">VHG-AN-0.1-800G</a>	Acid Number Standard 0.1 mgKOH/g	800g
<a href="#">VHG-AN-0.5-100G</a>	Acid Number (AN) Standard: 0.5 mg KOH/g in Hydrocarbon Oil	100g
<a href="#">VHG-AN-0.5-400G</a>	Acid Number Standard 0.5 mgKOH/g	400g
<a href="#">VHG-AN-0.5-800G</a>	Acid Number Standard 0.5 mgKOH/g	800g
<a href="#">VHG-AN-1.5-100G</a>	Acid Number (AN) Standard: 1.5 mg KOH/g in Hydrocarbon Oil	100g
<a href="#">VHG-AN-1.5-400G</a>	Acid Number Standard 1.5 mgKOH/g	400g
<a href="#">VHG-AN-1-100G</a>	Acid Number (AN) Standard: 1.0 mg KOH/g in Hydrocarbon Oil	100g
<a href="#">VHG-AN-1-400G</a>	Acid Number Standard 1 mgKOH/g	400g
<a href="#">VHG-AN-1-800G</a>	Acid Number Standard 1 mgKOH/g	800g
<a href="#">VHG-AN-2-100G</a>	Acid Number Standard 2 mgKOH/g	100g
<a href="#">VHG-AN-2-50G</a>	Acid Number (AN) Standard: 2.0 mg KOH/g in Hydrocarbon Oil	50g
<a href="#">VHG-AN-3-400G</a>	Acid Number Standard 3 mgKOH/g	400g
<a href="#">VHG-AN-3-50G</a>	Acid Number (AN) Standard: 3.0 mg KOH/g in Hydrocarbon Oil	50g
<a href="#">VHG-AN-3-800G</a>	Acid Number Standard 3 mgKOH/g	800g

## Section 4

# Sucrose Brix Reference Standards

Paragon's dual accredited ISO 17025 / ISO 17034 Sucrose Standards are for use in the calibration and verification of all types of refractometers e.g. handheld, Abbe and high accuracy digital instruments. Sucrose Brix Standards do not contain stabilisers and are manufactured from high purity materials in accordance with International Commission for Uniform Methods of Sugar Analysis (ICUMSA) methodology.

Paragon Scientific – Sucrose Brix/RI Reference Standards		
Product Number	Product Description	Size (ml)
<a href="#">ALK-SS00</a>	Pure Water (0.00 Brix / 1.33299 RI)	15ml
<a href="#">ALK-SS02</a>	Sucrose (2.00 °Brix / 1.33586 RI)	15ml
<a href="#">ALK-SS05</a>	Sucrose (5.00 °Brix / 1.34026 RI)	15ml
<a href="#">ALK-SS075</a>	Sucrose (7.50 °Brix / 1.34401 RI)	15ml
<a href="#">ALK-SS10</a>	Sucrose (10.00 °Brix / 1.34782 RI)	15ml
<a href="#">ALK-SS112</a>	Sucrose (11.20 °Brix / 1.34968 RI)	15ml
<a href="#">ALK-SS115</a>	Sucrose (11.50 °Brix / 1.35015 RI)	15ml
<a href="#">ALK-SS12</a>	Sucrose (12.00 °Brix / 1.35093 RI)	15ml
<a href="#">ALK-SS125</a>	Sucrose (12.50 °Brix / 1.35171 RI)	15ml
<a href="#">ALK-SS15</a>	Sucrose (15.00 °Brix / 1.35568 RI)	15ml
<a href="#">ALK-SS16</a>	Sucrose (16.00 °Brix / 1.35729 RI)	15ml
<a href="#">ALK-SS20</a>	Sucrose (20.00 °Brix / 1.36384 RI)	15ml
<a href="#">ALK-SS25</a>	Sucrose (25.00 °Brix / 1.37233 RI)	15ml
<a href="#">ALK-SS30</a>	Sucrose (30.00 °Brix / 1.38115 RI)	15ml
<a href="#">ALK-SS32</a>	Sucrose (32.00 °Brix / 1.38478 RI)	15ml
<a href="#">ALK-SS35</a>	Sucrose (35.00 °Brix / 1.39032 RI)	15ml
<a href="#">ALK-SS40</a>	Sucrose (40.00 °Brix / 1.39986 RI)	15ml
<a href="#">ALK-SS45</a>	Sucrose (45.00 °Brix / 1.40978 RI)	15ml
<a href="#">ALK-SS50</a>	Sucrose (50.00 °Brix / 1.42009 RI)	15ml
<a href="#">ALK-SS55</a>	Sucrose (55.00 °Brix / 1.43080 RI)	15ml
<a href="#">ALK-SS60</a>	Sucrose (60.00 °Brix / 1.44193 RI)	15ml

## Section 5

# Smoke Point Reference Standards

Paragon's dual accredited ISO 17025 & 17034 Smoke Point Reference Fuel Blends are ideal for the calibration and/ or verification of analytical instrumentation for automatic or manual measurement of smoke point using ASTM D1322 and IP 598. The range of Smoke Point Reference Fuel Blends corresponds to listed composition values as per Table 1 in ASTM D1322 and IP 598. We also offer a set kit for use with automatic apparatus, which includes x1 of each reference blend. Applications – Kerosene and aviation turbine fuel.

Paragon Scientific – Smoke Point Reference Standards		
Product Number	Product Description	Size (ml)
<a href="#">ALK-SPRF-1</a>	Smoke Point Reference Fuel Blend 1 for 14.7 mm (40/60 %v/v)	100ml
<a href="#">ALK-SPRF-2</a>	Smoke Point Reference Fuel Blend 2 for 20.2 mm (25/75 %v/v)	100ml
<a href="#">ALK-SPRF-3</a>	Smoke Point Reference Fuel Blend 3 for 22.7 mm (20/80 %v/v)	100ml
<a href="#">ALK-SPRF-4</a>	Smoke Point Reference Fuel Blend 4 for 25.8 mm (15/85 %v/v)	100ml
<a href="#">ALK-SPRF-5</a>	Smoke Point Reference Fuel Blend 5 for 30.2 mm (10/90 %v/v)	100ml
<a href="#">ALK-SPRF-6</a>	Smoke Point Reference Fuel Blend 6 for 35.4 mm (5/95 %v/v)	100ml
<a href="#">ALK-SPRF-7</a>	Smoke Point Reference Fuel Blend 7 for 42.8 mm (0/100 %v/v)	100ml
<a href="#">ALK-SPRF-KIT-7</a>	Smoke Point Fuel Blend Kit, (Blends 1 – 7: 14.7 mm, 20.2 mm, 22.7 mm, 25.8 mm, 30.2 mm, 35.4 mm and 42.8 mm)	7x100ml



## Section 6

# Relative Density Reference Standards

Paragon's dual accredited ISO 17025 / ISO 17034 Relative Density (Specific Gravity) standards are designed for the calibration or verification of instruments, used to measure density and relative density of materials at the desired test temperature within the range of 15 °C to 25 °C. These include, but are not limited to, automatic instruments and hydrometers. All density measurements are made in accordance with ASTM D1480, for density and relative density (specific gravity) of viscous materials by Bingham Pycnometer. The Relative density was calculated by dividing the density, as obtained from ASTM D1480, by the density of water at the reference temperature.

Paragon Scientific – Relative Density Reference Standards		
Product Number	Product Description	Size (ml)
<a href="#">ALK-RDEN15-01</a>	Relative Density Standard 15 °C, (Nominal value 0.6654 at 15 °C)	500ml
<a href="#">ALK-RDEN15-02</a>	Relative Density Standard 15 °C, (Nominal value 0.7183 at 15 °C)	500ml
<a href="#">ALK-RDEN15-03</a>	Relative Density Standard 15 °C, (Nominal value 0.7807 at 15 °C)	500ml
<a href="#">ALK-RDEN15-04</a>	Relative Density Standard 15 °C, (Nominal value 0.8111 at 15 °C)	500ml
<a href="#">ALK-RDEN15-05</a>	Relative Density Standard 15 °C, (Nominal value 0.8494 at 15 °C)	500ml
<a href="#">ALK-RDEN15-06</a>	Relative Density Standard 15 °C, (Nominal value 0.8682 at 15 °C)	500ml
<a href="#">ALK-RDEN15-07</a>	Relative Density Standard 15 °C, (Nominal value 0.8811 at 15 °C)	500ml
<a href="#">ALK-RDEN20-01</a>	Relative Density Standard 20 °C, (Nominal value 0.6609 at 20 °C)	500ml
<a href="#">ALK-RDEN20-02</a>	Relative Density Standard 20 °C, (Nominal value 0.7142 at 20 °C)	500ml
<a href="#">ALK-RDEN20-03</a>	Relative Density Standard 20 °C, (Nominal value 0.7769 at 20 °C)	500ml
<a href="#">ALK-RDEN20-04</a>	Relative Density Standard 20 °C, (Nominal value 0.8386 at 20 °C)	500ml
<a href="#">ALK-RDEN20-05</a>	Relative Density Standard 20 °C, (Nominal value 0.8452 at 20 °C)	500ml
<a href="#">ALK-RDEN20-06</a>	Relative Density Standard 20 °C, (Nominal value 0.8723 at 20 °C)	500ml
<a href="#">ALK-RDEN25-01</a>	Relative Density Standard 25 °C, (Nominal value 0.6564 at 25 °C)	500ml
<a href="#">ALK-RDEN25-02</a>	Relative Density Standard 25 °C, (Nominal value 0.7101 at 25 °C)	500ml
<a href="#">ALK-RDEN25-03</a>	Relative Density Standard 25 °C, (Nominal value 0.7730 at 25 °C)	500ml
<a href="#">ALK-RDEN25-04</a>	Relative Density Standard 25 °C, (Nominal value 0.8352 at 25 °C)	500ml
<a href="#">ALK-RDEN25-05</a>	Relative Density Standard 25 °C, (Nominal value 0.8693 at 25 °C)	500ml



## Section 7

# Multi-Parameter Reference Standards

Paragon's range of Multi-Parameter Certified Reference Materials certified to Refractive Index and Density, in accordance with our ISO 17025 and iso 17034 accreditations. These are available in four different matrices, Pure Water, Dodecane, Dichlorotoluene, and Bromonaphthalene. Each material is certified for Refractive Index and density at 15, 20 & 25 °C.

### Paragon Scientific – Multi-Parameter Refractive Index & Density Certified Reference Standards, at 15°C, 20°C and 25°C

Product Number	Product Description	Size (ml)
<a href="#">ALK-PS-RVD-01</a>	Multi-Parameter Refractive Index & Density Certified Reference Material (Nominal RI: 1.3330 @ 20 °C)	30ml
<a href="#">ALK-PS-RVD-02</a>	Multi-Parameter Refractive Index & Density Certified Reference Material (Nominal RI: 1.4217 @ 20 °C)	30ml
<a href="#">ALK-PS-RVD-03</a>	Multi-Parameter Refractive Index & Density Certified Reference Material (Nominal RI: 1.5463 @ 20 °C)	30ml
<a href="#">ALK-PS-RVD-04</a>	Multi-Parameter Refractive Index & Density Certified Reference Material Nominal RI: 1.6579 @ 20 °C)	30ml



## Section 8: Refractive Index Reference Standards

### Section 8

# Refractive Index Reference Standards

Paragon's dual accredited ISO 17025 / ISO 17034 Refractive Index Certified Reference Materials (CRMs) for the calibration and verification of temperature-controlled refractometers, with each CRM providing certified values for refractive index measurements at 20 °C, 25 °C and 30 °C.

Paragon Scientific – Refractive Index Certified Reference Standards at 20°C, 25°C and 30°C		
Product Number	Product Description	Size (ml)
<a href="#">ALK-PS-RI-01</a>	Refractive Index Certified Reference Material (Nominal RI Value 1.3325 at 25 °C)	10ml
<a href="#">ALK-PS-RI-01K</a>	Refractive Index Certified Reference Material (Nominal RI Value 1.3325 at 25 °C)	5x10ml
<a href="#">ALK-PS-RI-02</a>	Refractive Index Certified Reference Material (Nominal RI Value 1.3891 at 25 °C)	10ml
<a href="#">ALK-PS-RI-02K</a>	Refractive Index Certified Reference Material (Nominal RI Value 1.3891 at 25 °C)	5x10ml
<a href="#">ALK-PS-RI-03</a>	Refractive Index Certified Reference Material (Nominal RI Value 1.4023 at 25 °C)	10ml
<a href="#">ALK-PS-RI-03K</a>	Refractive Index Certified Reference Material (Nominal RI Value 1.4023 at 25 °C)	5x10ml
<a href="#">ALK-PS-RI-04</a>	Refractive Index Certified Reference Material (Nominal RI Value 1.4196 at 25 °C)	10ml
<a href="#">ALK-PS-RI-04K</a>	Refractive Index Certified Reference Material (Nominal RI Value 1.4196 at 25 °C)	5x10ml
<a href="#">ALK-PS-RI-05</a>	Refractive Index Certified Reference Material (Nominal RI Value 1.4206 at 25 °C)	10ml
<a href="#">ALK-PS-RI-05K</a>	Refractive Index Certified Reference Material (Nominal RI Value 1.4206 at 25 °C)	5x10ml
<a href="#">ALK-PS-RI-06</a>	Refractive Index Certified Reference Material (Nominal RI Value 1.4573 at 25 °C)	10ml
<a href="#">ALK-PS-RI-06K</a>	Refractive Index Certified Reference Material (Nominal RI Value 1.4573 at 25 °C)	5x10ml
<a href="#">ALK-PS-RI-07</a>	Refractive Index Certified Reference Material (Nominal RI Value 1.4941 at 25 °C)	10ml
<a href="#">ALK-PS-RI-07K</a>	Refractive Index Certified Reference Material (Nominal RI Value 1.4941 at 25 °C)	5x10ml
<a href="#">ALK-PS-RI-08</a>	Refractive Index Certified Reference Material (Nominal RI Value 1.5349 at 25 °C)	10ml

Continued on the next page

## Section 8: Refractive Index Reference Standards

Continued from previous page

Paragon Scientific – Refractive Index Certified Reference Standards at 20°C, 25°C and 30°C		
Product Number	Product Description	Size (ml)
<a href="#">ALK-PS-RI-08K</a>	Refractive Index Certified Reference Material (Nominal RI Value 1.5349 at 25 °C)	5x10ml
<a href="#">ALK-PS-RI-09</a>	Refractive Index Certified Reference Material (Nominal RI Value 1.5440 at 25 °C)	10ml
<a href="#">ALK-PS-RI-09K</a>	Refractive Index Certified Reference Material (Nominal RI Value 1.5440 at 25 °C)	5x10ml
<a href="#">ALK-PS-RI-10</a>	Refractive Index Certified Reference Material (Nominal RI Value 1.6556 at 25 °C)	10ml
<a href="#">ALK-PS-RI-10K</a>	Refractive Index Certified Reference Material (Nominal RI Value 1.6556 at 25 °C)	5x10ml

## Section 9

# Flash Point Reference Standards

### Flash Point Primary Certified Reference Materials

Manufactured and certified by a method specific inter-laboratory study in strict accordance with ISO 17034, using only those laboratories accredited to ISO 17025 for the test. Intended for use in the verification of flash point apparatus on at least an annual basis, as required by flash point methodology. Please see Section 12 Single Parameter Certified Reference Materials (CRMs and CRMUs) for these certified reference materials.

### Secondary Working Flash Point Reference Standards

Manufactured and certified in strict accordance with our ISO 17025 and ISO 17034 accreditations to ASTM D92 and ASTM D93, Procedure A. Intended for the regular verification of flash point apparatus.

Cleveland Open Cup Method Reference Standards		
Product Number	Product Description	Size (ml)
<a href="#">ALK-FP-COC-1</a>	Flash Point Reference Standard (Nominal value: 84 °C)	3x80ml
<a href="#">ALK-FP-COC-3</a>	Flash Point Reference Standard (Nominal value: 164 °C)	3x80ml
<a href="#">ALK-FP-COC-4</a>	Flash Point Reference Standard (Nominal value: 205 °C)	3x80ml
<a href="#">ALK-FP-COC-5</a>	Flash Point Reference Standard (Nominal value: 259 °C)	3x80ml
<a href="#">ALK-FP-COC-6</a>	Flash Point Reference Standard (Nominal value: 118 °C)	3x80ml

## Section 9: Flash Point Reference Standards

Pensky Martens Method Reference Standards		
Product Number	Product Description	Size (ml)
<a href="#">ALK-FP-PMCC-1</a>	Flash Point Reference Standard (Nominal value: 55.0 °C)	3x80ml
<a href="#">ALK-FP-PMCC-2</a>	Flash Point Reference Standard (Nominal value: 75.5 °C)	3x80ml
<a href="#">ALK-FP-PMCC-3</a>	Flash Point Reference Standard (Nominal value: 109.0 °C)	3x80ml
<a href="#">ALK-FP-PMCC-4</a>	Flash Point Reference Standard (Nominal value: 137.5 °C)	3x80ml
<a href="#">ALK-FP-PMCC-5</a>	Flash Point Reference Standard (Nominal value: 175.0 °C)	3x80ml
<a href="#">ALK-FP-PMCC-6</a>	Flash Point Reference Standard (Nominal value: 219.5 °C)	3x80ml

## Section 10

# Certified Ethanol Reference Standards

Dual accredited ISO 17025 & 17034 Certified Ethanol Standards are certified for % ABV (Alcohol by Volume) and are designed to be used for the calibration and/ or verification of alcoholmeters or densitometers used to determine alcohol content. The alcohol content is certified in accordance with the UK Revenue & Customs Laboratory Alcohol Table.

Applications – Food, Beverage and alcohol.

Paragon Scientific – Certified Ethanol Reference Standards		
Product Number	Product Description	Size (ml)
<a href="#">ALK-ETWA05</a>	Certified Ethanol Standard – 5 % ABV (Alcohol by Volume)	25ml
<a href="#">ALK-ETWA15</a>	Certified Ethanol Standard – 15 % ABV (Alcohol by Volume)	25ml
<a href="#">ALK-ETWA40</a>	Certified Ethanol Standard – 40 % ABV (Alcohol by Volume)	25ml
<a href="#">ALK-ETWA70</a>	Certified Ethanol Standard – 70 % ABV (Alcohol by Volume)	25ml

## Section 11: Density Reference Standards

### Section 11

# Density Reference Standards

Our dual accredited ISO 17025 & 17034 Density standards are available in temperature ranges between 15 – 150 °C, where the material is fluid at the desired test temperature, certified by ASTM D1480. Typically used in, but not limited to, density measuring equipment e.g., the oscillating u-tube method, such as ASTM D4052 / IP 365 and other internationally equivalent methodology.

Included in the density offering is a pure water density standard. Certified by ASTM D1480 at the temperatures of 15, 20 & 25 °C.

#### Paragon Scientific – Pure Water Density Reference Standards, Density Values at 15, 20 & 25 °C

Product Number	Product Description	Size (ml)
<a href="#">ALK-DEN-WAT</a>	Pure Water Density Standard (Nominal: 0.99909 g/mL @ 15 °C)	60ml
<a href="#">ALK-DEN-WAT3</a>	Pure Water Density Standard (Nominal: 0.99909 g/mL @ 15 °C)	3x60ml

#### Paragon Scientific – Density Reference Standards

Product Number	Product Description	Size (ml)
<a href="#">ALK-DEN100-01</a>	Density Standard 100 °C, (Nominal density value 0.7645 at 100 °C)	60ml
<a href="#">ALK-DEN100-02</a>	Density Standard 100 °C, (Nominal density value 0.8124 at 100 °C)	60ml
<a href="#">ALK-DEN100-03</a>	Density Standard 100 °C, (Nominal density value 0.8550 at 100 °C)	60ml
<a href="#">ALK-DEN15-01</a>	Density Standard 15 °C, (Nominal density value 0.6654 at 15 °C)	60ml
<a href="#">ALK-DEN15-02</a>	Density Standard 15 °C, (Nominal density value 0.7183 at 15 °C)	60ml
<a href="#">ALK-DEN15-03</a>	Density Standard 15 °C, (Nominal density value 0.7807 at 15 °C)	60ml
<a href="#">ALK-DEN15-04</a>	Density Standard 15 °C, (Nominal density value 0.8111 at 15 °C)	60ml
<a href="#">ALK-DEN15-05</a>	Density Standard 15 °C, (Nominal density value 0.8494 at 15 °C)	60ml
<a href="#">ALK-DEN15-06</a>	Density Standard 15 °C, (Nominal density value 0.8648 at 15 °C)	60ml
<a href="#">ALK-DEN15-07</a>	Density Standard 15 °C, (Nominal density value 0.8811 at 15 °C)	60ml
<a href="#">ALK-DEN15-08</a>	Density Standard 15 °C, (Nominal density value 0.9413 at 15 °C)	60ml
<a href="#">ALK-DEN15-09</a>	Density Standard 15 °C, (Nominal density value 0.9823 at 15 °C)	60ml
<a href="#">ALK-DEN15-10</a>	Density Standard 15 °C, (Nominal density value 1.0248 at 15 °C)	60ml
<a href="#">ALK-DEN15-11</a>	Density Standard 15 °C, (Nominal density value 1.0687 at 15 °C)	60ml

Continued on the next page

## Section 11: Density Reference Standards

Continued from previous page

Paragon Scientific – Density Reference Standards		
Product Number	Product Description	Size (ml)
<a href="#">ALK-DEN15-12</a>	Density Standard 15 °C, (Nominal density value 1.1280 at 15 °C)	60ml
<a href="#">ALK-DEN15-13</a>	Density Standard 15 °C, (Nominal density value 1.1962 at 15 °C)	60ml
<a href="#">ALK-DEN15-14</a>	Density Standard 15 °C, (Nominal density value 1.2829 at 15 °C)	60ml
<a href="#">ALK-DEN15-15</a>	Density Standard 15 °C, (Nominal density value 1.6300 at 15 °C)	60ml
<a href="#">ALK-DEN150-01</a>	Density Standard 150 °C, (Nominal density value 0.7288 at 150 °C)	60ml
<a href="#">ALK-DEN150-02</a>	Density Standard 150 °C, (Nominal density value 0.7816 at 150 °C)	60ml
<a href="#">ALK-DEN150-03</a>	Density Standard 150 °C, (Nominal density value 0.8287 at 150 °C)	60ml
<a href="#">ALK-DEN20-01</a>	Density Standard 20 °C, (Nominal density value 0.6609 at 20 °C)	60ml
<a href="#">ALK-DEN20-02</a>	Density Standard 20 °C, (Nominal density value 0.7142 at 20 °C)	60ml
<a href="#">ALK-DEN20-03</a>	Density Standard 20 °C, (Nominal density value 0.7769 at 20 °C)	60ml
<a href="#">ALK-DEN20-04</a>	Density Standard 20 °C, (Nominal density value 0.8386 at 20 °C)	60ml
<a href="#">ALK-DEN20-05</a>	Density Standard 20 °C, (Nominal density value 0.8452 at 20 °C)	60ml
<a href="#">ALK-DEN20-06</a>	Density Standard 20 °C, (Nominal density value 0.8723 at 20 °C)	60ml
<a href="#">ALK-DEN20-07</a>	Density Standard 20 °C, (Nominal density value 0.9378 at 20 °C)	60ml
<a href="#">ALK-DEN20-08</a>	Density Standard 20 °C, (Nominal density value 0.9811 at 20 °C)	60ml
<a href="#">ALK-DEN20-09</a>	Density Standard 20 °C, (Nominal density value 1.0236 at 20 °C)	60ml
<a href="#">ALK-DEN20-10</a>	Density Standard 20 °C, (Nominal density value 1.0669 at 20 °C)	60ml
<a href="#">ALK-DEN20-11</a>	Density Standard 20 °C, (Nominal density value 1.1256 at 20 °C)	60ml
<a href="#">ALK-DEN20-12</a>	Density Standard 20 °C, (Nominal density value 1.1915 at 20 °C)	60ml
<a href="#">ALK-DEN20-13</a>	Density Standard 20 °C, (Nominal density value 1.2800 at 20 °C)	60ml
<a href="#">ALK-DEN20-14</a>	Density Standard 20 °C, (Nominal density value 1.6218 at 20 °C)	60ml
<a href="#">ALK-DEN25-01</a>	Density Standard 25 °C, (Nominal density value 0.6564 at 25 °C)	60ml
<a href="#">ALK-DEN25-02</a>	Density Standard 25 °C, (Nominal density value 0.7101 at 25 °C)	60ml
<a href="#">ALK-DEN25-03</a>	Density Standard 25 °C, (Nominal density value 0.7730 at 25 °C)	60ml
<a href="#">ALK-DEN25-04</a>	Density Standard 25 °C, (Nominal density value 0.8352 at 25 °C)	60ml
<a href="#">ALK-DEN25-05</a>	Density Standard 25 °C, (Nominal density value 0.8693 at 25 °C)	60ml
<a href="#">ALK-DEN25-06</a>	Density Standard 25 °C, (Nominal density value 0.9342 at 25 °C)	60ml
<a href="#">ALK-DEN25-07</a>	Density Standard 25 °C, (Nominal density value 0.9797 at 25 °C)	60ml
<a href="#">ALK-DEN25-08</a>	Density Standard 25 °C, (Nominal density value 1.0222 at 25 °C)	60ml

Continued on the next page

## Section 11: Density Reference Standards

Continued from previous page

Paragon Scientific – Density Reference Standards		
Product Number	Product Description	Size (ml)
<a href="#">ALK-DEN25-09</a>	Density Standard 25 °C, (Nominal density value 1.0650 at 25 °C)	60ml
<a href="#">ALK-DEN25-10</a>	Density Standard 25 °C, (Nominal density value 1.1231 at 25 °C)	60ml
<a href="#">ALK-DEN25-11</a>	Density Standard 25 °C, (Nominal density value 1.1868 at 25 °C)	60ml
<a href="#">ALK-DEN25-12</a>	Density Standard 25 °C, (Nominal density value 1.2771 at 25 °C)	60ml
<a href="#">ALK-DEN25-13</a>	Density Standard 25 °C, (Nominal density value 1.6136 at 25 °C)	60ml
<a href="#">ALK-DEN40-01</a>	Density Standard 40 °C, (Nominal density value 0.6426 at 40 °C)	60ml
<a href="#">ALK-DEN40-02</a>	Density Standard 40 °C, (Nominal density value 0.6977 at 40 °C)	60ml
<a href="#">ALK-DEN40-03</a>	Density Standard 40 °C, (Nominal density value 0.7934 at 40 °C)	60ml
<a href="#">ALK-DEN40-04</a>	Density Standard 40 °C, (Nominal density value 0.8250 at 40 °C)	60ml
<a href="#">ALK-DEN40-05</a>	Density Standard 40 °C, (Nominal density value 0.8716 at 40 °C)	60ml
<a href="#">ALK-DEN50-01</a>	Density Standard 50 °C, (Nominal density value 0.7864 at 50 °C)	60ml
<a href="#">ALK-DEN50-02</a>	Density Standard 50 °C, (Nominal density value 0.8102 at 50 °C)	60ml
<a href="#">ALK-DEN50-03</a>	Density Standard 50 °C, (Nominal density value 0.8659 at 50 °C)	60ml
<a href="#">ALK-DEN60-01</a>	Density Standard 60 °C, (Nominal density value 0.7924 at 60 °C)	60ml
<a href="#">ALK-DEN60-02</a>	Density Standard 60 °C, (Nominal density value 0.8201 at 60 °C)	60ml
<a href="#">ALK-DEN60-03</a>	Density Standard 60 °C, (Nominal density value 0.8688 at 60 °C)	60ml
<a href="#">ALK-DEN80-01</a>	Density Standard 80 °C, (Nominal density value 0.7785 at 80 °C)	60ml
<a href="#">ALK-DEN80-02</a>	Density Standard 80 °C, (Nominal density value 0.8246 at 80 °C)	60ml
<a href="#">ALK-DEN80-03</a>	Density Standard 80 °C, (Nominal density value 0.8578 at 80 °C)	60ml

## Section 12

# Single Parameter Certified Reference Materials (CRMs and CRMUs)

We produce a variety of Single Parameter Certified Reference Materials, particularly aimed at the petroleum industry. Those materials produced under our ISO 17034 accreditation (denoted CRMU) are certified by a method specific inter-laboratory study using those laboratories that are accredited to ISO 17025 for the test. The materials certified outside our accreditation (denoted CRM) are certified by inter-laboratory study.

### Primary Certified Flash Point standards

Certified by a method specific inter-laboratory study using test laboratories that are accredited to ISO 17025. Certified in strict accordance with our ISO 17025 and ISO 17034 accreditations to ASTM D92 and ASTM D93, Procedure A.

Paragon Scientific – Certified Reference Material under UKAS accreditation standards		
Product Number	Product Description	Size (ml)
<a href="#">ALK-CRMU-ABKR</a>	Certified Reference Material Abel Flash point, Jet Aviation Fuel (Nominal value: 40.3 °C)	250ml
<a href="#">ALK-CRMU-ADKR</a>	Certified Reference Material Acidity (Jet), Jet Aviation Fuel (Nominal value: 0.0067 mg KOH/g)	250ml
<a href="#">ALK-CRMU-CFGO</a>	Certified Reference Material Cold Filter Plugging Point, Diesel (Nominal value: -21.7 °C)	250ml
<a href="#">ALK-CRMU-COC-HIGH</a>	Certified Reference Material, Cleveland Open Cup Flash Point (Nominal value: 262.0 °C)	3x80ml
<a href="#">ALK-CRMU-COC-LOW</a>	Certified Reference Material, Cleveland Open Cup Flash Point (Nominal value: 111.0 °C)	3x80ml
<a href="#">ALK-CRMU-COC-MID</a>	Certified Reference Material, Cleveland Open Cup Flash Point (Nominal value: 161.4 °C)	3x80ml
<a href="#">ALK-CRMU-CPGO</a>	Certified Reference Material Cloud Point Standard, Diesel (Nominal value: -7.7 °C)	250ml
<a href="#">ALK-CRMU-DEGA</a>	Certified Reference Material Density Standard, Gasoline (Nominal value 0.72587 g/mL @ 15 °C)	250ml
<a href="#">ALK-CRMU-DEGO</a>	Certified Reference Material Density Standard, Diesel (Nominal value 0.83418 g/mL @ 15 °C)	250ml
<a href="#">ALK-CRMU-DEKR</a>	Certified Reference Material Density Standard, Jet Aviation Fuel (Nominal value 0.79684 g/mL @ 15 °C)	250ml
<a href="#">ALK-CRMU-DELU</a>	Certified Reference Material Density Standard, Lubricant, (Nominal value 0.86709 g/mL @ 15 °C)	250ml
<a href="#">ALK-CRMU-DIGA</a>	Certified Reference Material Distillation Standard, Gasoline (Nominal values from 32.8 to 173.3 °C)	250ml

Continued on the next page

## Section 12: Single Parameter Certified Reference Materials (CRMs and CRMUs)

Continued from previous page

Paragon Scientific – Certified Reference Material under UKAS accreditation standards		
Product Number	Product Description	Size (ml)
<a href="#">ALK-CRMU-DIGO</a>	Certified Reference Material Distillation Standard, Diesel (Nominal values from 160.8 to 355 °C)	250ml
<a href="#">ALK-CRMU-DIKR</a>	Certified Reference Material Distillation Standard, Jet Aviation Fuel (Nominal value: 158.7 to 268.2 °C)	250ml
<a href="#">ALK-CRMU-FCLU</a>	Certified Reference Flash Point Standard – Cleveland, Lubricant (Nominal value: 257.5 °C)	250ml
<a href="#">ALK-CRMU-FIKR</a>	Certified Reference Material FIA Aromatics, Jet Aviation Fuel (Nominal value: 17.17%)	250ml
<a href="#">ALK-CRMU-FRKR</a>	Certified Reference Material Freezing Point, Jet Aviation Fuel (Nominal value: -53.7 °C)	250ml
<a href="#">ALK-CRMU-MPGO</a>	Multi-Parameter Certified Reference Material, Diesel	500ml
<a href="#">ALK-CRMU-MPLU</a>	Multi-Parameter Certified Reference Material, Lubricant	500ml
<a href="#">ALK-CRMU-PMCC-HIGH</a>	Certified Reference Material, Pensky Martens Flash Point (Nominal value: 210.5 °C)	3x80ml
<a href="#">ALK-CRMU-PMCC-LOW</a>	Certified Reference Material, Pensky Martens Flash Point (Nominal value: 76.5 °C)	3x80ml
<a href="#">ALK-CRMU-PMCC-MID</a>	Certified Reference Material, Pensky Martens Flash Point (Nominal value: 135.4 °C)	3x80ml
<a href="#">ALK-CRMU-PMGO</a>	Certified Reference Flash Point Standard Pensky Martens, Diesel (Nominal value: 66.1 °C)	250ml
<a href="#">ALK-CRMU-PMLU</a>	Certified Reference Flash Point Standard - PMCC Procedure B, Lubricant (Nominal value: 190.5 °C)	250ml
<a href="#">ALK-CRMU-PMLUB</a>	Certified Reference Flash Point Standard - PMCC Procedure B, Lubricant (Nominal value: 100.7 °C)	250ml
<a href="#">ALK-CRMU-PPGO</a>	Certified Reference Pour Point Standard, Diesel (Nominal value: -34.0 °C)	250ml
<a href="#">ALK-CRMU-PPLU</a>	Certified Reference Pour Point Standard, Lubricant (Nominal value: -11.2 °C)	250ml
<a href="#">ALK-CRMU-PPLU1</a>	Certified Reference Pour Point Standard, Lubricant (Nominal value: -26.1 °C)	250ml
<a href="#">ALK-CRMU-PPLU2</a>	Certified Reference Pour Point Standard, Lubricant (Nominal value: -38.4 °C)	250ml
<a href="#">ALK-CRMU-SPKR</a>	Certified Reference Smoke Point - Automatic Certified Reference Material, Jet Aviation Fuel (Nominal: 23.76 mm)	250ml
<a href="#">ALK-CRMU-SUKR</a>	Certified Reference Mercaptan Sulphur Standard in Jet Aviation Fuel (approx 7.3 mg/kg)	250ml
<a href="#">ALK-CRMU-TAKR</a>	Certified Reference Material TAG Flash Point, Jet Aviation Fuel (Nominal value: 40.4° C)	250ml



**Section 12:** Single Parameter Certified Reference Materials (CRMs and CRMUs)

Paragon Scientific – Certified Reference Materials		
Product Number	Product Description	Size (ml)
<a href="#">ALK-CRM-ABKR</a>	Certified Reference Material Abel Flash point, Jet Aviation Fuel (Nominal value: 40.3 °C)	250ml
<a href="#">ALK-CRM-ACGA</a>	Certified Reference Material Aromatics Content, Gasoline (Nominal value: 27.1%)	250ml
<a href="#">ALK-CRM-ADKR</a>	Certified Reference Material Acidity (Jet), Jet Aviation Fuel (Nominal value: 0.0085 mg KOH/g)	250ml
<a href="#">ALK-CRM-APKR</a>	Certified Reference Material Aniline Point, Jet Aviation Fuel (Nominal value: 56.60 °C)	250ml
<a href="#">ALK-CRM-BEGA</a>	Certified Reference Material Benzene Content (Nominal value 0.62% Volume)	250ml
<a href="#">ALK-CRM-CFGO</a>	Certified Reference Material Cold Filter Plugging Point, Diesel (Nominal value: -16.3 °C)	250ml
<a href="#">ALK-CRM-CNGO</a>	Certified Reference Material Cetane Number, Diesel (Nominal value: 52.6)	1l
<a href="#">ALK-CRM-DEGA</a>	Certified Reference Material Density Standard, Gasoline (Nominal value: 0.7429 g/mL @ 15 °C)	250ml
<a href="#">ALK-CRM-DEGO</a>	Certified Reference Material Density Standard, Diesel (Nominal value 0.8375 g/mL @ 15 °C)	250ml
<a href="#">ALK-CRM-DEKR</a>	Certified Reference Material Density Standard, Jet Aviation Fuel (Nominal value 0.8025 g/mL @ 15 °C)	250ml
<a href="#">ALK-CRM-DIGA</a>	Certified Reference Material Distillation Standard, Unleaded Gasoline (Nominal value: 37.4 to 181.3 °C)	250ml
<a href="#">ALK-CRM-DIGO</a>	Certified Reference Material Distillation Standard, Diesel (Nominal values from 172.8 to 364.5 °C)	250ml
<a href="#">ALK-CRM-FAGO</a>	Certified Reference Material Fatty Acid Methyl Ester (FAME) Standard, Diesel (Nominal value 6.5%)	250ml
<a href="#">ALK-CRM-FIKR</a>	Certified Reference Material FIA Aromatics, Jet Aviation Fuel (Nominal value: 20.1% Volume)	250ml
<a href="#">ALK-CRM-OMGA</a>	Certified Reference Material Motor Octane Number, Gasoline (Nominal value: 85.9)	1l
<a href="#">ALK-CRM-ORGA</a>	Certified Reference Material Research Octane Number, Gasoline (Nominal value: 97.2)	1l
<a href="#">ALK-CRM-SUKR</a>	Certified Reference Material Mercaptan Sulphur in Hydrocarbons, Jet Aviation Fuel (Nominal value: 15.2 mg/kg)	250ml
<a href="#">ALK-CRM-TAKR</a>	Certified Reference Flash Point Standard, TAG, Jet Aviation Fuel (Nominal value: 41.4 °C)	250ml
<a href="#">ALK-CRM-VPGA</a>	Certified Reference Material Reid Vapour Pressure Standard, Gasoline (Nominal value: 58.9 kPa)	250ml

**Section 12:** Single Parameter Certified Reference Materials (CRMs and CRMUs)



## Section 13

# Conductivity Reference Standards

Here's our range of ISO 17025 & ISO 17034 accredited Conductivity Standards for calibration and verification of various conductivity meters. Conductivity is a physical analysis measuring the capability of a liquid to pass current ( $\mu\text{S}/\text{cm}$ ). Our conductivity standards range from 2 to 500,000  $\mu\text{S}/\text{cm}$ . Measurement is done by an in-house method based on ASTM D1125, Standard Test Methods for Electrical Conductivity and Resistivity of Water, however usage is applicable to most conductivity applications.

Conductivity Reference Standards		
Product Number	Product Description	Size (ml)
<a href="#">ALK-COND0002-25-125ML</a>	Conductivity standard, 2 microSiemens/cm at 25 °C, NIST Traceable, ACS	125ml
<a href="#">ALK-COND0005-25-125ML</a>	Conductivity standard, 5 microSiemens/cm at 25 °C, NIST Traceable, ACS	125ml
<a href="#">ALK-COND0010-25-125ML</a>	Conductivity standard, 10 microSiemens/cm at 25 °C, NIST Traceable, ACS	125ml
<a href="#">ALK-COND0100-25-500ML</a>	Conductivity standard, 100 microSiemens/cm at 25 °C, NIST Traceable, ACS	500ml
<a href="#">ALK-COND010K-25-500ML</a>	Conductivity standard, 10,000 microSiemens/cm at 25 °C, NIST Traceable, ACS	500ml
<a href="#">ALK-COND012K-25-500ML</a>	Conductivity standard, 12,880 microSiemens/cm at 25 °C, NIST Traceable, ACS	500ml
<a href="#">ALK-COND0147-25-500ML</a>	Conductivity standard, 147 microSiemens/cm at 25 °C, NIST Traceable, ACS	500ml
<a href="#">ALK-COND0500-25-500ML</a>	Conductivity standard, 500 microSiemens/cm at 25 °C, NIST Traceable, ACS	500ml
<a href="#">ALK-COND1000-25-500ML</a>	Conductivity standard, 1000 microSiemens/cm at 25 °C, NIST Traceable, ACS	500ml
<a href="#">ALK-COND100K-25-500ML</a>	Conductivity standard, 100,000 microSiemens/cm at 25 °C, NIST Traceable, ACS	500ml
<a href="#">ALK-COND1413-25-500ML</a>	Conductivity standard, 1413 microSiemens/cm at 25 °C, NIST Traceable, ACS	500ml
<a href="#">ALK-COND500K-25-500ML</a>	Conductivity standard, 500,000 microSiemens/cm at 25 °C, NIST Traceable, ACS	500ml
<a href="#">VHG-CONDNA100-1L</a>	Conductivity @ 100 $\mu\text{mho}/\text{cm}$ in H <sub>2</sub> O	1l
<a href="#">VHG-CONDNA1K-1L</a>	Conductivity @ 1000 $\mu\text{mho}/\text{cm}$ in H <sub>2</sub> O	1l

## Section 14

# Colour Reference Standards

A wide range of colour standards for the calibration and verification of colour measuring instruments. The current range includes materials accredited under ISO 17025 / ISO 17034, as well as materials certified under the ISO 9001 quality system.

Those supplied under ISO 17025 / ISO 17034 include ASTM Colour Standards (ASTM D1500, ASTM D6045), Gardner Colour Standards (ASTM D1544, ASTM D166) and Saybolt Colour Standards (ASTM D156, ASTM D6045).

Paragon Scientific – Colour Reference Standards – ASTM Method		
Product Number	Product Description	Size (ml)
<a href="#">ALK-ASTM10</a>	ASTM 1, Colour	500ml
<a href="#">ALK-ASTM30</a>	ASTM 3, Colour	500ml
<a href="#">ALK-ASTM50</a>	ASTM 5, Colour	500ml
<a href="#">ALK-ASTM70</a>	ASTM 7, Colour	500ml
<a href="#">ALK-ASTM05</a>	ASTM <0.5, Colour	500ml

Paragon Scientific – Colour Reference Standards – Saybolt Method		
Product Number	Product Description	Size (ml)
<a href="#">ALK-SAYB08</a>	Saybolt +25, Colour	500ml
<a href="#">ALK-SAYB07</a>	Saybolt +12, Colour	500ml
<a href="#">ALK-SAYB06</a>	Saybolt 0, Colour	500ml
<a href="#">ALK-SAYB02</a>	Saybolt -10, Colour	500ml

## Section 14: Colour Reference Standards

Paragon Scientific – Colour Reference Standards – Lovibond RYBN Method		
Product Number	Product Description	Size (ml)
<a href="#">ALK-134080.00</a>	Colour Reference Standard Lovibond RYBN Colour 0.8R 2.0Y 0.1N (5¼")	500ml
<a href="#">ALK-134090</a>	Colour Reference Standard Lovibond RYBN Colour 1.4R 4.0Y 0.5N (5¼")	500ml
<a href="#">ALK-134100</a>	Colour Reference Standard Lovibond RYBN Colour 2.0R 7.0Y 0.5N (5¼")	500ml
<a href="#">ALK-134110</a>	Colour Reference Standard Lovibond RYBN Colour 2.1R 11.0Y 0.5N (5¼")	500ml
<a href="#">ALK-134120</a>	Colour Reference Standard Lovibond RYBN Colour 2.5R 14.0Y 0.7N (5¼")	500ml
<a href="#">ALK-134130.00</a>	Colour Reference Standard Lovibond RYBN Colour 3.1R 22.0Y 0.85N (5¼")	500ml
<a href="#">ALK-134230</a>	Colour Reference Standard Lovibond RYBN Colour 3.4R 30.0Y 0.9N (5¼")	500ml

Paragon Scientific – Colour Reference Standards – Pt-Co/Hazen/APHA		
Product Number	Product Description	Size (ml)
<a href="#">ALK-133991</a>	Colour Reference Standard Pt-Co/Hazen/APHA Colour, Nominal Certified Value 0	500ml
<a href="#">ALK-134140</a>	Colour Reference Standard Pt-Co/Hazen/APHA Colour, Nominal Certified Value 5	500ml
<a href="#">ALK-134150</a>	Colour Reference Standard Pt-Co/Hazen/APHA Colour, Nominal Certified Value 10	500ml
<a href="#">ALK-134160</a>	Colour Reference Standard Pt-Co/Hazen/APHA Colour, Nominal Certified Value 15	500ml
<a href="#">ALK-134170</a>	Colour Reference Standard Pt-Co/Hazen/APHA Colour, Nominal Certified Value 30	500ml
<a href="#">ALK-134180</a>	Colour Reference Standard Pt-Co/Hazen/APHA Colour, Nominal Certified Value 50	500ml
<a href="#">ALK-134190</a>	Colour Reference Standard Pt-Co/Hazen/APHA Colour, Nominal Certified Value 100	500ml
<a href="#">ALK-462803</a>	Colour Reference Standard Pt-Co/Hazen/APHA Colour, Nominal Certified Value 500	500ml

## Section 14: Colour Reference Standards

### Paragon Scientific – Colour Reference Standards – Gardner Method

Product Number	Product Description	Size (ml)
<a href="#">ALK-GARD02</a>	Gardner Value 2, Colour	500ml
<a href="#">ALK-GARD05</a>	Gardner Value 5, Colour	500ml
<a href="#">ALK-GARD08</a>	Gardner Value 8, Colour	500ml

### Paragon Scientific – Colour Reference Standards – AOCS-Tintometer Method

Product Number	Product Description	Size (ml)
<a href="#">ALK-134240</a>	Colour Reference Standard AOCS-Tintometer Colour 0.4R 2.0Y (5¼")	500ml
<a href="#">ALK-134250</a>	Colour Reference Standard AOCS-Tintometer Colour 1.6R 9.0Y (5¼")	500ml
<a href="#">ALK-134260</a>	Colour Reference Standard AOCS-Tintometer Colour 1.9R 12Y (5¼")	500ml
<a href="#">ALK-134270</a>	Colour Reference Standard AOCS-Tintometer Colour 2.5R 20Y (5¼")	500ml
<a href="#">ALK-134280.00</a>	Colour Reference Standard AOCS-Tintometer Colour 3.0R 28Y (5¼")	500ml



## Section 15

# Simulated Distillation Reference Standards

Our Simulated Distillation standards for use with ASTM D3552 and D7169. These are standard test methods for boiling point distribution of petroleum samples by gas chromatography. Our Simulated distillation standards must pass rigorous quality control in our ISO/IEC 17025-certified laboratory.

### VHG – Simulated Distillation Reference Standards

Product Number	Product Description	Size (ml)
<a href="#">VHG-POLYW-1000-1ML</a>	Simulated Distillation Reference Material for C5-C120, 1mL	1ml

## Section 16

# Red Eye in Diesel Reference Standards

ASTM D6258 Red Dye #26 Standards for determination of Solvent Red 164 Dye Concentration in Diesel Fuels to ensure your diesel fuel dye testing accuracy. Many commercially available reagents specified in ASTM D6258 have a significant problem with impurities. This can cause testing labs and final data end users unintended issues, up to and including serious penalties if their diesel fuel is found out of compliance. Our VHG is proud to solve these challenges with our accurate and thoroughly characterized dye testing standard for use with ASTM D6258. The calibration and control standards are produced from a highly characterized, certified dye, delivering consistent lot-to-lot performance for each calibration standard with excellent correlation statistics. The product is created using our stringent manufacturing processes, which are accredited to ISO 17034 and certified to ISO 9001 and was prepared to the certified concentrations shown on the reverse side by gravimetric methods in accordance to ASTM D6258. Each lot of the product passes rigorous Quality Control by UV-VIS in our laboratory accredited to ISO/IEC 17025.

### VHG – Red Dye in Diesel Reference Standards

Product Number	Product Description	Size (ml)
<a href="#">VHG-DSLRED26-QC-100</a>	Red Dye #26 QC Check 10 mg/L	100ml
<a href="#">VHG-DSLREDDYE-KIT-6X100</a>	Red Dye #26 Calibration Kit: 0, 3, 6, 9, 12, 15 mg/L 6x100 mL	6x100ml

## Section 17

# Soot in Diesel Reference Standards

Performing regular diesel soot check scan help you save time and money by limiting waste from unnecessary oil changes, while protecting and extending the life of your diesel engines. LGC Industrial provides standards you can rely on for use in accordance with ASTM Methods D5967, D7686, and D7844. Our Soot Content Standards help streamline your lab operations by eliminating in-house soot standard preparation while strengthening the integrity of your testing results. We prepare these standards to the certified concentrations shown by Thermal Gravimetric Analysis (TGA) according to ASTM D5967, Appendix A4. Each lot of the product passes rigorous Quality Control by TGA. Each product is shipped with a comprehensive Certificate of Analysis (CoA). We ensures the accuracy of this standard for 24 months from the certification date.

VHG – Soot in Diesel Reference Standards		
Product Number	Product Description	Size (ml)
<a href="#">VHG-SOOT-A-50</a>	0.5-2 wt% Soot in Diesel Engine Oil	50mL
<a href="#">VHG-SOOT-B-50</a>	2-4 wt% Soot in Diesel Engine Oil	50mL
<a href="#">VHG-SOOT-BLK-50</a>	0 wt% Soot in Diesel Engine Oil	50mL
<a href="#">VHG-SOOT-C-50</a>	4-6 wt% Soot in Diesel Engine Oil	50mL
<a href="#">VHG-SOOT-D-50</a>	6-9 wt% Soot in Diesel Engine Oil	50mL
<a href="#">VHG-SOOT-E-50</a>	9-12 wt% Soot in Diesel Engine Oil	50mL
<a href="#">VHG-SOOT-SET</a>	Soot Content Standard Set (contains one of each of the following: SOOT-BLK, SOOT-A, SOOT-B, SOOT-C, SOOT-D, SOOT-E)	6x50mL





## Section 18

# Moisture Content Reference Standards

Our Moisture Content Standards are intended for use as a certified reference material in the determination of water in motor oil by Karl Fischer Crackle Test or Karl Fischer Titration.

### Crackle Test Reference Standards

These standards are intended for use in the detection of water in motor oil by crackle test. It must be well shaken prior to use. This standard is NOT intended for use with Karl Fischer titration methods. Our standards are manufactured and certified under a quality control system that is accredited to both ISO 9001 and ISO/IEC 17025. This standard was prepared to the nominal concentration using gravimetric methods. Tools: The balances used in the preparation of VHG standards are calibrated regularly with traceability to NIST.

### Titration Reference Standards

Our Karl Fischer titration standards are intended for use in accordance with ASTM Method D6304. These CRMs were manufactured and certified under a quality management system that is accredited to ISO 9001, ISO 17034 and ISO/IEC 17025. This CRM was prepared to the nominal water concentration of 0.5% (w/w) using gravimetric methods. The balances used in the preparation of our CRMs are calibrated regularly with traceability to NIST. The certified concentration was determined by using coulometric Karl Fischer titration, in accordance with ASTM D6304 Procedure C, and employing a water evaporator accessory.

#### VHG – Crackle Test Reference Standards

Product Number	Product Description	Size (ml)
<a href="#">VHG-CTR-0.1P-100</a>	Crackle Test Reference Standard: 0.1 wt% H <sub>2</sub> O in 10W30 Motor Oil	100ml
<a href="#">VHG-CTR-0.5P-100</a>	Crackle Test Reference Standard: 0.5 wt% H <sub>2</sub> O in 10W30 Motor Oil	100ml
<a href="#">VHG-CTR-1.0P-100</a>	Crackle Test Reference Standard: 1.0 wt% H <sub>2</sub> O in 10W30 Motor Oil	100ml
<a href="#">VHG-CTR-BLK-100</a>	Crackle Test Reference Standard: 0 wt% H <sub>2</sub> O in 10W30 Motor Oil	100ml

#### VHG – Titration Reference Standards

Product Number	Product Description	Size (ml)
<a href="#">VHG-KF-0.05P-100</a>	Karl Fischer Standard: 0.05 wt% H <sub>2</sub> O in 10W30 Motor Oil	100ml
<a href="#">VHG-KF-0.1P-100</a>	Karl Fischer Standard: 0.1 wt% H <sub>2</sub> O in 10W30 Motor Oil	100ml
<a href="#">VHG-KF-0.5P-100</a>	Karl Fischer Standard: 0.5 wt% H <sub>2</sub> O in 10W30 Motor Oil	100ml
<a href="#">VHG-KF-1.0P-100</a>	Karl Fischer Standard: 1.0 wt% H <sub>2</sub> O in 10W30 Motor Oil	100ml

## Section 19

# Particle Count Reference Standards

Parti-Count Particle Count Standards for predictive engine failure analysis. Analysis of particles in fluids may be the most important way to monitor the condition of your engines. Once wear begins, the rate of wear usually increases rapidly, and studying trends in your particle count data can yield surprising dividends. VHG Parti-Count Particle Count Standards, compliant with ISO 11171 and traceable to NIST SRM 2806b, are cost-effective calibration standards for the verification of automatic particle counters. Parti-Count is created using our stringent manufacturing processes, which are accredited to ISO 17034 and certified to ISO 9001. This solution contains ISO Medium Test Dust (MTD) sourced directly from NIST RM 8631a for use with Automatic Particle Counters (APC) calibrated to ISO 11171. Each lot of Parti-Count passes rigorous Quality Control by Automatic Particle Counter (APC) in our laboratory accredited to ISO/IEC 17025.

VHG – Particle Count Reference Standards – Parti-Count™ Particle Count Verification Fluid		
Product Number	Product Description	Size (ml)
<a href="#">VHG-PCMTD-5-125</a>	Parti-Count™ Particle Count Verification Fluid: 5 mg/L ISO MTD. 4, 6, 10, 14, 18, 21, 38, 50, and 70µm channels reported	125ml
<a href="#">VHG-PCMTD-5-500</a>	Parti-Count™ Particle Count Verification Fluid: 5 mg/L ISO MTD. 4, 6, 10, 14, 18, 21, 38, 50, and 70µm channels reported	500ml



**1 USA + Canada + Mexico \***  
Tel: +1 (603) 935 4100  
Email: industrial@lgcgroup.com

**2 United Kingdom \***  
Paragon Scientific  
Tel: +44 (0)20 8943 8480  
Email: sales@paragon-sci.com

**3 France**  
Tel: +33 (0)3 88 04 82 82  
Email: fr@lgcgroup.com

**4 Germany**  
Tel: +49 (0)281 9887 0  
Email: de@lgcgroup.com

**5 Italy**  
Tel: +39 02 22476412  
Email: it@lgcgroup.com

**6 Middle East**  
Tel: +49 (0)281 9887 0  
Email: global.sales@lgcgroup.com

**7 Nordic countries**  
Tel: +49 (0)281 9887 0  
Email: de@lgcgroup.com

**8 Poland**  
Tel: +48 22 751 31 40  
Email: pl@lgcgroup.com

**9 South Africa \***  
Tel: +27 (0)11 466 4321  
Email: sales.za@lgcgroup.com

**10 Spain**  
Tel: +34 (0)93 308 4181  
Email: es@lgcgroup.com

**11 China**  
Tel: +86 400 9216156  
Email: info.china@lgcgroup.com

**12 India**  
Tel: +91 (0)90 8297 4025  
Email: india@lgcgroup.com

**Distributor Network**  
Tel: +49 (0) 281 9887 250  
Email: global.sales@lgcgroup.com

We leverage an extensive distributor network where there is no sales office represented.

\* Indicates a Centre of Excellence

For full listing of offices and distributors please see [lgcstandards.com](http://lgcstandards.com)



Paragon Scientific  Ltd

**Industrial**  
VHG | ARMI | MBH